COOKE INTERNATIONAL SERVICES
TEST & MEASUREMENT INSTRUMENTS
Unit 4, Fordingbridge Site, Main Road
Bersham, Bognor Regis, West Sussex PO22 0EB
Telephone: (0243) 545111/2 Fax: (0243) 542457

B221A
Q221A
AA221

INSTRUCTION MANUAL

WAYNE KERR

UNIVERSAL BRIDGE
LOW IMPEDANCE ADAPTOR
AUTOBALANCE ADAPTOR

UNIVERSAL BRIDGE B221A

LOW IMPEDANCE ADAPTOR Q221A

AUTOBALANCE ADAPTOR AA221

The Reference Number of this Publication is TP2/4

THE WAYNE KERR COMPANY LIMITED

NEW MALDEN · SURREY · ENGLAND

Telephone : ~~01-942 2202~~

01-399-6751

CONTENTS (Part 1 - B221A)

	Page
BRIDGE SET, UNIVERSAL, CT530	6
INTRODUCTION	7
SPECIFICATION	9
OPERATING INSTRUCTIONS	10
Setting Up	10
Measurement Procedure	10
Interpretation of Results	11
Source	11
Detector	11
Test Jigs	12
Lead Corrections	12
Major and Minor Terms	13
Ceramic Capacitors	13
Electrolytic Capacitors	14
Loss Angle	14
Iron-Cored Inductors	14
Inductors Passing D.C.	14
Screened Components	15
Temperature Coefficients	16
In situ Measurements	16
Network Characteristics	17
Attenuator Measurements	17
Effectiveness of Transformer Screens	18
Turns Ratios	18
50 Hz – 20 kHz Operation	18
PRINCIPLE OF OPERATION	19
CIRCUIT DESCRIPTION	20
Power Supply	20
Source	20
Detector	20
Bridge Circuit	20
Standards	21
Trimming of Standards	22
Residual Capacitance Trimming	22
Main Bridge Trimming	22
Unknown Connexions and Range Switch	23
Measurement Voltage	23

(*Continued overleaf*)

Page

MAINTENANCE AND SERVICE 24
Source and Detector 24
Overall Gain 24
Oscillator Frequency 24
Bridge Accuracy 24
Checking and Adjusting Standards 25
Conductance Standards 25
Capacitance Standards 26
Phase Correction 27

B221A COMPONENTS LIST 49

RECIPROCAL TABLES 52
Reproduced from Tables of Mathematical Functions by J. B. Dale (Edward Arnold Limited) by permission of the publishers.

CONTENTS (Part 2 - Q221A)

INTRODUCTION 28

SPECIFICATION 28

OPERATING INSTRUCTIONS 29
Measurement Procedure 29
Interpretation of Results 30
Measurement of Reactance 30
Connexion to the Unknown 30
Errors in the Minor Component 31
Measurement of R.F. Coils 31
Measurement of Mutual Inductance 31
Polarizing Electrolytic Capacitors 32
Accumulator Resistance 32
50 Hz – 20 kHz Operation 32
Discrimination and Accuracy 33
Current through the Unknown 33

PRINCIPLE OF OPERATION 34

CIRCUIT DESCRIPTION 34

MAINTENANCE AND SERVICE 34

Q221A COMPONENTS LIST 48

PARALLEL TO SERIES CONVERSION 37

CONTENTS (Part 3 - AA221)

	Page
INTRODUCTION	39
SPECIFICATION	40
OPERATING INSTRUCTIONS – AA221 with B221A only	41
Setting Up	41
Measurement Procedure	41
Measurement Conditions	42
Monitoring Outputs	42
OPERATING INSTRUCTIONS – AA221 with B221A & Q221A	43
Setting Up	43
Measurement Procedure	43
Measurement Conditions	44
SPECIALISED MEASUREMENTS	44
POWER SUPPLY UNIT PS109	44
PRINCIPLE OF OPERATION	45
CIRCUIT DESCRIPTION	45
Source	45
Amplifier	45
Feedback Network	46
Phase-Sensitive Detector (G)	46
90° Circuit	46
Phase-Sensitive Detector (C)	46
Voltmeter/Record Outputs	46
MAINTENANCE AND SERVICE	47
Source	47
Amplifier/Detector	47
Voltmeter Output	47
AA221 COMPONENTS LIST	51

ILLUSTRATIONS (Part 1 - B221A)

		page
Frontispiece (Bridge B221A)		8
Fig. 1	Polarizing Ceramic Capacitors	13
Fig. 2	Polarizing Electrolytic Capacitors	14
Fig. 3	Inductors Passing D.C.	14
Fig. 4	Three-terminal Capacitor	15
Fig. 5	Three-terminal Network	16
Fig. 6	Three-terminal Measurements	16
Fig. 7	In Situ Measurements	16

		Page
Fig. 8	Shunting Effect	16
Fig. 9	Network Characteristics	17
Fig. 10	Attenuator Measurements	17
Fig. 11	Transformer Screens	18
Fig. 12	Principle of Operation	19
Fig. 13	Source Circuit	20
Fig. 14	Detector Circuit	20
Fig. 15	Bridge Circuit	21
Fig. 16	Third G Decade	21
Fig. 17	C Phase Trimming	22
Fig. 18	Residual Capacitance	22
Fig. 19	Trim Controls	22
Fig. 20	Range Switch	23
B221A	Circuit Diagram	48
Table 1	Maximum Permissible Current Through Bridge	15
Table 2	Transformer Parameters	17
Table 3	Range Switch Connexions	17
Table 4	Effect of frequency on *C* max. and *L* min.	18
Table 5	Voltage Readings	24

ILLUSTRATIONS (Part 2 - Q221A)

	Low Impedance Adaptor	28
Fig. 21	Connexion to the Unknown	29
Fig. 22	Adaptor Range Plate	30
Fig. 23	Measurement of Mutual Inductance	31
Fig. 24	Polarizing Electrolytic Capacitors	32
Fig. 25	Principle of Operation	34
	Capacitance Conversion Chart	35
	Correction Curves (Resistance)	36
	Q221A Circuit Diagram	48

ILLUSTRATIONS (Part 3 - AA221)

	Autobalance Adaptor	38
Fig. 26	PS109 Circuit Diagram	44
Fig. 27	AA221 Block Diagram	45
	AA221 Circuit Diagram	50

BRIDGE SET, UNIVERSAL, CT530
10S/6625-99-948-8768
and
BRIDGE SET, UNIVERSAL, CT530 SET
6625-99-955-3163

The Services version of Wayne Kerr Universal Bridge B221A is designated B221S and has the Service reference number CT530. Performance of the two instruments is identical but, in addition to certain internal constructional differences which do not affect the information contained in this Manual, there are certain external differences.

1 CT530 is supplied with a lid suitable for housing all the measurement cables and the power input cable. The lid can be attached to the front panel for transit purposes or to the rear panel whilst the instrument is in operation.

2 Power input to the CT350 is by means of a Plessey Mk. IV connector on the front panel in place of the moulded three-pin connector which is fitted to the rear panel of the Commercial instrument.

3 In place of the illuminated push on/push off button for a.c. power, the CT530 has a toggle switch. The two magic eyes provide a visual reminder that the instrument is switched 'on'.

4 The CT530 does not include Low Impedance Adaptor Q221A.

5 The CT530 SET includes the Services version of Q221A (Extension Impedance, Bridge Universal, Q221S), 6625-99-955-3161, and instrument packing cases.

Part 1 - Universal Bridge B221A

INTRODUCTION

Universal Bridge B221A operates at audio frequencies and provides accurate values for all types of electrical components over an extremely wide measurement range. The internal Standards are of conductance and capacitance, but the Bridge measures positive or negative conductance or resistance, together with capacitance or inductance. Two displays provide, simultaneously but without any interaction, values for the resistive and reactive terms of any Unknown, including all cyphers, the decimal point and the units of measurement.

The Bridge has an internal source and detector operating at a fixed frequency (normally 1592 Hz to simplify calculations involving $2\pi f$) but can be operated with an external source and detector at any audio frequency. Suitable instruments for this purpose include the Wayne Kerr Audio Signal Generator S121 and (above 100 Hz) Waveform Analyser A321. Full details of these instruments are available separately: their use with the B221A Bridge is described in this Manual.

A valuable property of the Bridge is its ability to make accurate measurements even when long connecting cables are used and whilst the component under test remains connected in a circuit. Thus, remote measurements can be undertaken when environmental tests are to be made on components situated in ovens, refrigerators and pressure or humidity enclosures. Further, individual components forming part of a printed circuit board or encapsulated assembly can be measured without disconnecting them from the associated wiring and components. This facility is achieved by using the Neutral terminals of the Bridge, enabling any arm of a delta-network to be measured in the presence of the two other arms.

This (the fourth) issue of the Instruction Manual has been revised to present a more straightforward operating procedure, and the Manual is now produced in three parts. Part 1 deals with the Universal Bridge and Part 2 with the Low Impedance Adaptor Q221A, these two parts each including the relevant information on operation with an external source and detector. Part 3 (omitted from certain issues) contains full information on Autobalance Adaptor AA221 for reference purposes. Autobalance Bridge B641 serves as replacement for the B221A/AA221 combination. A completely separate publication is available for the B641, also for the various Cells and Adaptors which simplify use of the Bridge for measurements of conductivity or permittivity.

WAYNE KERR
Universal Bridge B221
3 6·3
µMho
4
·0 0 4 1 8
µF
Range
G
C
E
Supply
Sensitivity
Trim G
Trim C

SPECIFICATION

Measurement Ranges

R (+ve or –ve)			G (+ve or –ve)			Range	C (×1)*			L (1592 Hz)‡		
Min (0·1%)	Max for 0·1%	First Divsn	First Divsn	Min for 0·1%	Max (0·1%)		First Divsn	Min for 0·1%	Max (0·1%)	Min (0·1%)§	Max for 0·1%§	First Divsn
9MΩ	100MΩ	50,000MΩ	·00002μmho	·01μmho	·111μmho	1	·002μμF	1μμF	11·1μμF	900H	10kH	5MH
0·9MΩ	10MΩ	5,000MΩ	·0002μmho	·1μmho	1·11μmho	2	·02μμF	10μμF	111μμF	90H	1kH	0·5MH
90kΩ	1MΩ	500MΩ	·002μmho	1μmho	11·1μmho	3	·2μμF	100μμF	1111μμF	9H	100H	50kH
9kΩ	100kΩ	50MΩ	·02μmho	10μmho	111μmho	4	·000002μF	·001μF	·0111μF	0·9H	10H	5kH
900Ω	10kΩ	5MΩ	·2μmho	100μmho	1111μmho	5	·00002μF	·01μF	·111μF	90mH	1H	500H
90Ω	1kΩ	500kΩ	·002mmho	1mmho	11·1mmho	6	·0002μF	·1μF	1·11μF	9mH	·1H	50H
9Ω	100Ω	50kΩ	·02mmho	10mmho	111mmho	7	·002μF	1μF	11·1μF	0·9mH	10mH	5H

Accuracy

±0·1% (see Table).
On Ranges 6 and 7, corrections must be made for the leads.

Discrimination

G and C: 0·02% of Maximum (see Table).
R and L: 0·02% of Minimum (see Table).

Source Frequency

Internal: 1592 Hz ($\omega = 10^4$) ±1%.
(Other frequencies to special order).
External: 50 Hz – 20 kHz.

Power Requirement

100 – 125V or 200 – 250V, 50 – 60 Hz.
Consumption approximately 25W.

Dimensions

Width: 17 in. (43 cm.).
Height: 11½ in. (29 cm.).
Depth: 7½ in. (19 cm.).

Weight

25 lb. (11 kg.).

The extension of the measurement ranges obtained with Adaptor Q221A is specified in Part 2.

* All figures become one-tenth of values quoted when C switch is set to 0·1. (Not available from AA221.)

‡ Frequency-dependent: $L = 1/\omega C$.
§ Accuracy of 0·1% applies if source frequency is measured: otherwise 2% applies.

OPERATING INSTRUCTIONS

SETTING UP

Power Requirement

Set the voltage selector at the rear of the instrument to the appropriate tapping for the supply voltage. Connect the power cable to a suitable plug [green to ground (earth), red to live and black or blue to neutral]. The instrument will operate from supplies of 100-125V and 200-250V, 50-60 Hz, and consumes approximately 25 watts. The Bridge is switched on or off by pushing the Supply button on the front panel.

Measurement Cables

The Bridge is supplied with two screened measurement cables, each terminated in a pair of crocodile clip leads. One cable (E) is associated with the voltage transformer on the source side of the Bridge: the other cable (I) is associated with the current transformer on the detector side of the Bridge circuit. In each case the green lead is connected to the screening (braid) and the leads are the two Neutral connections. The exposed metal sleeve between the two moulded sections is also connected to Neutral.

If Low Capacity Clips are to be used, these should be connected to the E and I sockets of the Bridge in place of the normal measurement cables. Again the Neutral connections are provided by the green crocodile clip leads.

Trimming

Link together the two Neutral leads. This can be achieved by clipping the green lead from one cable on to the green clip lead or metal sleeve of the other cable. Do not make any connection other than this. Turn the Range switch to position 7 (the number appears between the two magic eyes) and set all six decade controls to zero. Set the G and C switches to '1' and turn the Sensitivity control to an approximate mid-position. Adjust the Trim G and Trim C controls for maximum shadows on the magic eyes, finally increasing the sensitivity to maximum (fully clockwise). This trimming adjustment will hold good for all Ranges from 7 down to 3.

If the Range switch is turned to position 2 or 1, the trimming operation should be repeated on this Range. When either of these two Ranges is used, maximum accuracy will be obtained by substituting Low Capacity Clip Cables for the normal measurement cables. Whichever type of connection is employed, it is important (on Ranges 1 and 2) to arrange that the measurement leads are laid out for the trimming operation in the same relative positions as they will occupy for the actual measurement.

MEASUREMENT PROCEDURE

Connections to Unknown

For normal (*two-terminal*) measurements the Neutrals are linked as described under 'Trimming' and the component under test is connected between the two red clip leads. When the Low Capacitance Clips are used instead of the normal measurement cables, the Neutrals must be linked by connecting together the two green clips leads. The component under test is then connected between the two spring-loaded end connectors.

For *three-terminal* measurements the two green leads must remain linked and one of these is taken also to the third terminal of the Unknown. The Bridge then measures only the component situated between the two red leads.

Four-terminal measurements are made by separating the two Neutral connections. Input to the component or network under test is then provided by the red and green leads associated with the E cable and the output is applied to the red and green leads associated with the I cable.

Coarse Balance

Set the G switch to '1' for all measurements of positive conductance (or positive resistance). Set the C switch to '1' for capacitance measurements or to '–1' for inductance measurements. The Range switch should remain in position 7.

Adjust the Sensitivity control until the least sensitive (top left) magic eye shadow just begins to open. Rotate the continuously-variable (vernier) G and/or C controls until the magic eyes are seen to open. If this occurs when the controls are moved only slightly from the 0 setting it indicates that the Range switch setting must be reduced from position 7. In this instance the procedure could be repeated on Range 6 or 5.

If (on Range 7) the eyes have not opened when the vernier control(s) are fully clockwise, additional G or C must be inserted by operating the appropriate switched decade controls.

Fine Balance

When, on any range, a first indication of the value of the Unknown has been obtained on the vernier control(s), the Range switch setting can be reduced by two positions and the first and second digits of the value set-up on the appropriate switched decade controls. Re-adjustment of the vernier control(s) for maximum shadow will then provide the final balance with a four-figure reading of the major term. As the final balance is approached the Sensitivity control should be advanced progressively until fully clockwise.

Refer to final paragraph of 'Trimming' if the Range switch is moved to positions 1 or 2.

When the approximate value of any component is known before measurement, the appropriate Range and decade settings can be selected in advance. Much of the balance procedure described above can then be omitted.

INTERPRETATION OF RESULTS

When the Bridge is balanced, the two dial displays show values for the Unknown in terms of the equivalent *parallel* components of conductance and capacitance. When inductance is measured, the C multiplier switch will be in the '–1' position and the value of the Unknown is obtained from the C dial, using the expression:

$$L=1/(\omega^2 C)$$

where $\omega=2\pi \times$ frequency of measurement.

If the Bridge is operated at the normal frequency of 1592 *Hz*, $\omega^2=10^8$.

The equivalent *series* components of the Unknown can be derived from the Bridge dial readings at balance by using the following expressions:

$$R_s=1/[G_p(1+Q^2)] \qquad (1)$$

$$C_s=C_p(1+1/Q^2) \qquad (2)$$

$$L_s=1/[\omega^2 C_p(1+1/Q^2)] \qquad (3)$$

The value of Q can be calculated directly from the numerical readings presented by the decade controls without reference to the Range in use, the units of measurement or the position of the decimal point. At 1592 Hz, Q is always equal to the ratio of the numerical reading on the C decades to that on the G decades. For example, if balance is obtained on Range 4 with ·000625 μF and 12·50 μmhos (where the digits underlined are those presented by the decade controls—first C decade at zero in this example) then, ignoring the decimal point, Q = 0625/1250 = 0·5. If the C multiplier switch is on 0·1, the reading on the C decade must be divided by ten before Q is computed.

Note: When the Bridge is operated at 1592 Hz, $\omega^2=10^8$. Referring to equation (1), it can be seen that

$$\text{if } Q \ll 1,\ R_s \simeq 1/G_p.$$

Referring to equations (2) and (3), it can be seen that

$$\text{if } Q \gg 1,\ C_s \simeq C_p$$

$$\text{and } L_s \simeq 1/(\omega^2 C_p).$$

G_p and C_p are the Bridge readings of the equivalent parallel components.

Tables of reciprocals are provided at the end of this Manual.

SOURCE

The internal source operates at 1592 Hz ±1 per cent. This value was chosen to simplify calculations, since $2\pi f$ (or ω) is 10^4.

When it is desired to operate the Bridge at frequencies other than 1592 Hz, an external source can be employed. This should be capable of providing an output of 10-30V r.m.s. into an impedance of approximately 20kΩ. The larger output is required at low frequencies but the input to the Bridge must never exceed 40V r.m.s. The frequency coverage available depends on the measurement accuracy desired and the following figures provide a guide to this.

200Hz–10kHz: Better than ±0·25%

100Hz–200Hz and 10kHz–15kHz: Better than ±0·5%

50Hz–100Hz and 15kHz–20kHz: Better than ±1%

The source output should have a low harmonic content and *any d.c. component must be blocked externally*. Wayne Kerr Audio Signal Generator S121, which covers 10Hz to 120kHz, is an ideal instrument for this function. The external source is connected to the Bridge by means of a screened cable, using the jack and socket provided on the rear panel. Insertion of the plug automatically disconnects the internal source from the Bridge circuits.

DETECTOR

The internal detector is a two-stage amplifier tuned to 1592 Hz. One magic eye indicator is connected to a point between the two stages and the second magic eye is connected to the output of the second stage. Each magic eye has two shadows, of differing sensitivity. Thus, four degrees of sensitivity are available at any one time and a front-panel control enables the overall sensitivity to be varied. When the Bridge is balanced, the shadows are at a maximum (i.e. the eyes are open).

When the Bridge is operated with an external source at frequencies other than 1592 Hz, an external detector must also be provided. This must operate satisfactorily from an input falling to 10–20μV near balance and should have an input impedance of not less than 100kΩ. It is essential to employ a tuned amplifier, adjusted to the same frequency as the source, as this minimizes the effect of harmonics masking the point of balance at the required frequency.

When measurements are made at the lowest frequency (50–200 Hz) the sensitivity required is 1 to 5 microvolts. In general it is preferable to use a high-gain detector to obtain the required sensitivity, rather than to increase the source voltage. Wayne Kerr Waveform Analyser A321 (covering 20 Hz to 20 kHz) is suitable for use as

an external detector above 100 Hz. Connection to the external detector must be made with screened cable, using the jack and socket provided on the rear panel. Insertion of the plug automatically disconnects the internal detector from the Bridge circuits.

Note: When an external source and detector are in use, the Bridge should be disconnected from the a.c. power supply.

Certain bridges must be modified slightly before they are used with an external source and detector. The instruments affected are all serial numbers from 1293 to 1440 inclusive, except 1366, 1383 and 1428. Also affected are the following: 1443, 1444, 1445, 1446, 1448, 1456, 1460, 1462, 1463 and 1466. The modification needed on these instruments is as follows:

(1) Remove both jacks from the rear panel and expose the wiring.

(2) Strap together the two jack contacts nearest to the rear panel i.e. two metal braids. Repeat on the other jack.

(3) Re-assemble with the aluminium screens on to the rear panel.

TEST JIGS

It is possible, within the limits of the trimming controls, to balance out the conductance and stray capacitance of test jigs, etc. An accurate absolute measurement of the component under test can therefore be made. With the test jig alone connected to the measurement leads the Trim controls should be adjusted on the range to be used for measurement. If the component is now connected to the jig, its correct value can be measured directly.

LEAD CORRECTIONS

1 When the bridge is used for the measurement of conductances exceeding 10 millimhos (resistances of less than 100 ohms), capacitances exceeding 1 microfarad* or inductances of less than 10 millihenrys*, an allowance must be made for the series resistance and inductance of the measurement cables and the bridge transformers.

2 The simplest method for making the correction is dependent on the availability of a second component whose value is within 2% of the value of the Unknown. The bridge is trimmed and the Unknown measured normally, the conductance and capacitance values being noted (G1 and C1). A further measurement is made with the Unknown remaining in position, but the second component connected between the red and green leads of the E cable. New readings (G2 and C2) are noted.

3 The second component is then transferred to the red and green leads of the I cable (the Unknown remaining across the two red leads) and readings again noted (G3 and C3). The effect of the second component is given by adding the two variations from the original, i.e. (G1−G2)+(G1−G3) and (C1−C2) + (C1−C3). The original reading is corrected by adding the appropriate sum.

4 An alternative method relies only on the availability of two 10Ω resistors matched to within 2%. However, although the result is general, a calculation is necessary with each new measurement in order to apply the correction. The sequence of operations is similar to the previous method:

(i) Measure one 10Ω resistor, noting G1 and −C1.

(ii) Add second 10Ω (matched to 2%) between red and green leads of E cable. Rebalance. Note G2 and −C2.

(iii) Leaving first resistor across red-red, transfer second resistor to red and green of I cable. Rebalance. Note G3 and −C3.

5 The differences that occur on adding the second resistor are a measure of the series resistance and inductance of the leads and transformers.

6 *Resistance* When measuring 10 ohms on Range 7, the G vernier represents 0-100 milliohms series resistance. Thus the difference between conductance readings × 100 will give the series resistance directly in milliohms.

7 Total series resistance = [(G1−G2) + (G1−G3)] × 100 mΩ

To apply the correction, the measured value of the Unknown should first be converted to the series form (see 'Parallel to Series Conversions') and the total series lead and transformer resistance subtracted.

8 *Inductance* When measuring 10 ohms on Range 7, the C vernier represents 0−10 microhenrys series inductance. Thus the difference between capacitance readings × 100 will give the series inductance directly in microhenrys.

9 Total series inductance = {[(−C1) − (−C2)] + [(−C1) − (−C3)]} × 100μH

To apply the correction, the measured value of the unknown should first be converted to a series inductance and the total series lead and transformer inductance subtracted for inductance measurements. For capacitive unknown measurements on Range 7, the capacitance value measured must be converted to a series capacitive reactance, and the inductive reactance of the leads and transformers added. The true series or parallel capacitance can then be calculated.

* That is, reactance less than 100 ohms. Values quoted apply at 1592 Hz and must be computed for other measurement frequencies. The effect of shunt loading is described on page 16 ('In Situ Measurements').

NOTE:

$$L_s \text{ (true)} = \frac{1}{\omega^2 Cp\,(1 + 1/Q^2)} - L_T$$

For Capacitance X_s (true) =

$$\frac{1}{\omega Cp\,(1 + 1/Q^2)} - \omega L_T$$

$$C_s \text{ (true)} = \frac{1}{\omega X_s \text{ (true)}}$$

where L_T is the total series inductance of leads and transformers.

10 *Corrections for Range* 6

The method for obtaining the series resistance of leads and transformers is exactly the same as described for Range 7, but in this case a 100Ω resistor ± 2% or better is measured with 10Ω resistor loading. As before, the full vernier scales represent 100 mΩ for G and 10μH for C, but using the expressions

Series Resistance in milli-ohms = Total change in reading (mMho) × 1000

Series Inductance in microhenrys = Total change in reading (μF) × 1000

Worked Example for Range 7

1 Measurement of 10Ω resistor:
G1 = 97.65 mMho; −C1 = −0.040 μF.

2 Measurement of E cable loaded:
G2 = 96.39 mMho; −C2 = −0.050 μF

3 Measurement of I cable loaded:
G3 = 96.47 mMho; −C3 = −0.046 μF.

Total series resistance
= [(97.65 − 96.39) + (97.65 − 96.47)] × 100 mΩ
= (1.26 + 1.18) × 100 mΩ
= (2.44) × 100 mΩ
= 244 mΩ

Total series inductance
= [(−0.04 + 0.05) + (−0.04 + 0.046)] × 100 μH
= (0.01 + 0.006) × 100 μH.
= 1.6 μH.

Applying the correction to the 10Ω resistor measured

$$R_s = \frac{1}{G1} = \frac{1}{97.65 \times 10^{-3}} = 10.241\ \Omega$$

R_s (true) = 10.241 −0.244 Ω
= 9.997 Ω

MAJOR AND MINOR TERMS

The Range Switch operates on G and C simultaneously and the maximum values on each Range are such that

$$G_{max} = \omega C_{max}$$

(or $R_{min} = 1/\omega C_{max}$).

On any given range the larger of the two readings (G and C) is the major term of the Unknown and the smaller reading is the minor term. It will be realised that when the Range switch setting has been progressively reduced until a four-figure reading is obtained for the major term, the minor term may show as less than four significant figures. For many applications this is no limitation since the reduced reading accuracy is associated with the term which, in itself, is usually of less importance. However, it is possible to obtain a capacitance range 1/10th of normal by setting the C multiplier switch to the 0·1 position. Use of this facility provides increased reading accuracy for the minor term when this is the reactive component. When the C switch is set to 0·1, the Bridge dial reading of capacitance must be *divided* by 10.

CERAMIC CAPACITORS

With certain capacitors the value changes considerably with the value of d.c. polarizing voltage. In such instances a polarizing potential can be applied as shown in Fig. 1. The blocking capacitor should have a value at least 100 times that of the Unknown.



Fig. 1. Polarizing Ceramic Capacitors

ELECTROLYTIC CAPACITORS

The alternating potential applied across capacitors being measured on the Bridge is so low that it is generally unnecessary to polarize them. If it is necessary to measure them with a d.c. potential applied, the arrangement shown in Fig. 2 can be



Fig. 2. Polarizing Electrolytic Capacitors

used. The two capacitors, C, should be reasonably matched and of known value, very much smaller than C_u. The Unknown capacitance is given by

$$C_u = C^2/C_m$$

where C_m is the value measured on the Bridge. The phase-shaft introduced by this circuit may necessitate setting the G multiplier switch to —1 when balancing out the losses. The series loss resistance is given by the expression

$$r = -G_m/\omega^2 C^2.$$

LOSS ANGLE

The loss angle or power factor of capacitors can be computed from the values measured on the Bridge.

$$tan\ \delta = G/\omega C$$

$$cos\ \varphi = G/Y \simeq G/\omega C = tan\ \delta$$

An allowance for the impedance of the measuring leads must be made when dealing with large-value capacitors. This is described on page 12.

IRON-CORED INDUCTORS

Great care must be taken when interpreting the results obtained from measurements on iron-cored coils and transformer windings. With power transformers, for example, the core will have only its initial permeability due to the low value of the a.c. potential applied from the Bridge. The effective permeability at full operating voltage may be as much as twenty times higher.

With high-permeability materials such as mu-metal, the effective permeability changes so rapidly with excitation, even at very low levels of magnetization, that widely different values of inductance may be obtained on different ranges of the Bridge or even on two Bridges of the same pattern. Therefore, unless the operating conditions are simulated, the measurements made are of value only for comparison purposes.

INDUCTORS PASSING D.C.

It is often necessary to measure inductors carrying a known value of direct current and three ways of doing this are shown in Fig. 3. In Fig. 3(a) the d.c. supply, with a large-value capacitor by-passing it, is connected in series with the inductor being measured. This arrangement



Fig. 3. Inductors Passing D.C.

cannot be used, however, if one terminal of the d.c. supply is connected to Ground, as this would short-circuit the Bridge voltage transformer. The maximum values of current that may be passed through the Bridge are given in Table 1. Under no circumstances must these values be exceeded.

With the arrangement shown in Fig. 3(b) the d.c. supply is connected between the two Neutral leads. If one terminal of the d.c. supply is grounded it must be connected to the Neutral of the *E* lead. The limitations on current specified in Table 1 must be observed.

If the reactance of the by-pass capacitor shown in Fig. 3(a) and (b) is significant, it can be allowed for by calculating its equivalent inductance value at the Bridge frequency and *adding* this to the effective *series* value derived from the measurement.

TABLE 1

MAXIMUM PERMISSIBLE CURRENT THROUGH BRIDGE

Range	1	2	3	4	5	6	7
Max. D.C.	50mA	50mA	0·5A	0·5A	1·0A	1·0A	1·0A

Where it is required to pass a current through the inductor in excess of the values given in Table 1, the arrangement shown in Fig. 3(c) may be used. The resistors *R* must have a value of 50 to 100 times the impedance of the Unknown, and the capacitors *C* a reactance which is low compared with the value of *R*. The *T*-network may then be regarded as equivalent to a capacitor in parallel with a resistor connected straight across the Bridge. The equivalent *series* components of the Unknown are given by:

Series Resistance, $R_u = R^2.G_m$

Series Inductance, $L_u = R^2.C_m$

where G_m and C_m are the values measured on the Bridge.

SCREENED COMPONENTS

When resistors or capacitors are enclosed in metal cases and one terminal of the component is connected to the case, this terminal should be connected to the *E* lead from the Bridge. This method of connexion ensures minimum stray pick-up by the case and the component can then be measured in the usual manner.

When measuring sub-standard three-terminal capacitors, care must be taken to ensure that the conditions of measurement are the same as those under which the capacitor was calibrated. A three-terminal capacitor is shown schematically in Fig. 4. In addition to the main capacitance



Fig. 4. Three-terminal Capacitor

between the high- and low-potential terminals (*H* and *L*) there is the stray capacitance between each terminal and the metal case, *E*. Calibration is normally carried out with the low-potential side joined to the case. Under these conditions the stray capacitance between the *L* terminal and the case is shorted out, leaving the capacitance between *H* and the case in parallel with the main capacitance between *H* and *L*.

If the sub-standard is connected to the Bridge with the case connected to Neutral, *both* stray capacitances in the sub-standard will be neutralized and the reading obtained will be low. To this reading must be added the capacitance between *H* and the case. This can be measured on the Bridge by connecting the case to the voltage (*E*) lead from the Bridge, the *H* terminal to the current (*I*) lead and the *L* terminal to Neutral.

TEMPERATURE COEFFICIENTS

Since leads of any length may be used without their capacitance affecting the accuracy of the reactance measurement, and as lead resistance can be allowed for (see page 12), the Bridge is ideally suited to the measurement of temperature and other coefficients of components placed in an oven, refrigerator, pressure chamber, humidity chamber, etc. If full use is made of the decade controls, the discrimination of balance is 0·01 per cent of maximum for the range in use.

IN SITU MEASUREMENTS

It is often convenient to be able to measure an impedance *in situ*, without disconnecting any other components which may be associated with it. This applies particularly when components form part of a printed circuit or encapsulated assembly. Moreover, with certain test jigs, it is often impossible to 'disconnect' the effective shunt and stray capacitances. Generally speaking, any circuit can be resolved into a three-terminal network, the arrangement being as shown in Fig. 5.



Fig. 5. Three-terminal Network

The impedance to be measured is Z_u and the effect of impedances E–N and I–N must be removed. The arrangement can be considered as a π-network and is shown in Fig. 6 in this form, connected to the Bridge. Z_{EN} shunts the voltage transformer and Z_{IN} shunts the current transformer. In many instances these shunting effects may be disregarded (for reasons given under 'Principle of Operation'). In Fig. 7, for example, it may be possible to measure the impedance between A and B without disconnecting the other components simply by connecting point C to the Bridge Neutral.



Fig. 6. Three-terminal Measurements



Fig. 7. In Situ Measurements



Fig. 8. Shunting Effect

Important: *Under no circumstances must the inner of either measurement lead be connected to Ground while the green wire of the Bridge power cable is grounded.* The Neutral circuit of the Bridge is grounded internally and such action would short-circuit both Bridge transformers.

When it is necessary to connect either measurement lead (inner of E or I) to the chassis or framework of the equipment under test, this equipment must first be isolated from Ground. If this is not practicable, the Bridge must be isolated from Ground by disconnecting the green wire of the power cable from the Ground pin of the plug. Care must be exercised while operating the Bridge in this condition and the Ground connexion should be replaced when isolation is no longer necessary.

When the shunting effects (Z_{EN} and Z_{IN} of Fig. 6) cannot be ignored, the series resistance and leakage inductance of the Bridge transformers must be taken into account. The measurement circuit must now be drawn as in Fig. 8 and a simple calculation, described in succeeding paragraphs, must be made in order to correct the readings obtained from the bridge.

The loading effect of the Unknown impedance itself is small and can be ignored. It is only when the shunt impedances are low compared with Z_u that errors arising are of importance. Therefore Z_u can be considered as an open-circuit and the

shunting effects on either side of the Unknown can be considered separately.

The fall in potential at E caused by the load Z_{EN} will depend on the ratio of Z_{EN} to the impedance of R_e and L_e, the effective series resistance and leakage inductance of the voltage transformer. A similar argument applies to the current side of the Bridge and the two effects are additive.

Table 2 gives the average values of effective resistance and leakage inductance of the windings for each tapping. When it is necessary to correct for shunt loading the following formulae may be used to calculate the true value of impedance, Z_u, from the value read on the Bridge, Z_m.

With loading on the voltage transformer only:

$$Z_u=(1-Z_e/Z_{EN}).Z_m$$

where $Z_e=R_e+j\omega L_e$ ($\omega=10^4$).

With loading on the current transformer only:

$$Z_u=(1-Z_i/Z_{IN}).Z_m$$

where $Z_i=R_i+j\omega L_i$.

With both transformers loaded:

$$Z_u=[1-(Z_e/Z_{EN}+Z_i/Z_{IN})].Z_m.$$

TABLE 2

Turns (Tap)	Voltage Transformer (T1)			Current Transformer (T2)		
	R_e	L_e	X_e	R_i	L_i	X_i
1	0·025Ω	0·36μH	0·0036Ω	0·017Ω	0·63μH	0·0063Ω
10	0·025Ω	0·36μH	0·0036Ω	0·032Ω	0·9μH	0·009Ω
100	1·2Ω	18·0μH	0·18Ω	1·54Ω	46·0μH	0·46Ω
1000	72·0Ω	1·44mH	14·4Ω	68·0Ω	1·02mH	10·2Ω

If the loading is at the end of leads, add 75·6 milliohms and 0·3μH per yard of Uniradio 32. The transformer taps in use for the various ranges are given in Table 3.

TABLE 3

RANGE	TRANSFORMER TURNS	
	Voltage (T1)	Current (T2)
1	1000	1000
2	100	1000
3	100	100
4	10	100
5	10	10
6	1	10
7	1	1

NETWORK CHARACTERISTICS

The facility for three-terminal measurements and the readiness with which either or both the conductance and reactance Standards can be made effectively negative merely by switching to a winding of reverse sense, make the Bridge a most efficient instrument for measuring the characteristics of networks. Transfer admittance, for example, is measured simply by the arrangement



Fig. 9. Network Characteristics

shown in Fig. 9. The input of the network is connected between the voltage terminal and Neutral. The output is connected, in series with its terminating resistance R_T, between the I lead and Neutral. The right-hand side of R_T will be at Neutral potential when the Bridge is balanced and the network will then be correctly terminated.

Let the transfer admittance, defined as the current flowing in the terminating resistor for unit input voltage, be $|Y|\angle\theta$. Then, at balance:

$$|Y|\angle\theta=\pm G_m\pm j\omega C_m$$

where G_m and C_m are the values of conductance and capacitance read on the Bridge.

Then $|Y|=(G_m^2+\omega^2C_m^2)^{\frac{1}{2}}$

and $\theta=tan^{-1}(\pm\omega C_m/G_m)$.

ATTENUATOR MEASUREMENTS

The characteristics of attenuators can be measured quite simply on the Bridge and the circuit arrangement is shown in Fig. 10. With the resistor R_T in the position shown the attenuator is correctly terminated at balance, when the right-hand side of R_T will be at Neutral potential.



Fig. 10. Attenuator Measurements

If the Bridge is balanced with the attenuator set to zero, a value equal to the matching impedance R_T will be measured. If the attenuator steps are now switched in, the voltage attenuation can be calculated accurately from the ratio of the apparent change in value of R_T.

EFFECTIVENESS OF TRANSFORMER SCREENS

The effectiveness of screens between transformer windings can be determined using the arrangements shown in Fig. 11. The method of measurement when the transformer has a single inter-winding screen is shown in Fig. 11 (a) and, in Fig. 11 (b), the method of connexion when each winding has a separate screen is shown.



Fig. 11. Transformer Screens

With perfect screening there should be no feed through the transformers if connected as shown. Any effective capacitance between the windings will be indicated on the Bridge.

TURNS RATIOS

The turns ratio of a.f. transformers can be obtained by using an arrangement similar to that shown in Fig. 9, substituting the transformer for the Network. The value of R_T must be high compared with the output impedance of the transformer. Assuming the transformer has a primary: secondary turns ratio of $1:N$ and that the primary is connected to the E lead of the Bridge, the voltage produced across the secondary will be $N.E$. The secondary current (at balance) is given by $I = NE/R_T$. The conductance reading of the bridge, G_b, is equal to I/E. Therefore the turns ratio, N, is given by $R_T \times G_b$.

50Hz - 20kHz OPERATION

Reference should be made also to the sections on 'Source' and 'Detector', which include details of suitable external equipments and a guide to the measurement accuracy over the frequency range.

The Bridge is direct-reading in terms of conductance and capacitance, at all frequencies. However, when inductance is measured, conversion from the bridge capacitance dials will depend on the measurement frequency:

$$L = 1/(\omega^2 C) \simeq 1/(4\pi^2 f^2 C)$$
$$\text{mH} \simeq 25{\cdot}3/(f^2.\mu\text{F})$$

Permissible Bridge loading restricts the maximum capacitance that can be measured accurately at frequencies above 1592 Hz. [However, it should be noted that as the frequency is increased, the restriction in the Bridge coverage is accompanied by a corresponding extension in the measurement range of the Low Impedance Adaptor].

Table 4 provides a guide to the maximum capacitance and minimum inductance values that can be measured accurately on the Bridge at selected frequencies throughout the band. In all instances lead corrections must be applied when the admittance of the Unknown exceeds 10 millimhos (i.e. when the impedance is less than 100 ohms)—see 'Lead Corrections' (page 12).

When the Bridge is used on Range 1 or 2 (for the measurement of large values of resistance and/or reactance) at frequencies above 2000 Hz, particular care must be taken to avoid heavy shunt loading to the Neutral from either the voltage or current leads. If such loading is present, the measured value of the Unknown should be corrected using the formulae given on page 17.

TABLE 4

Source Frequency	Maximum Capacitance	Minimum Inductance
50Hz	10μF	1H
60Hz	10μF	0·7H
100Hz	10μF	250mH
200Hz	10μF	63mH
400Hz	10μF	16mH
800Hz	10μF	4mH
1000Hz	10μF	2·5mH
1592Hz	10μF	1mH
2kHz	8μF	0·8mH
4kHz	4μF	0·4mH
8kHz	2μF	0·2mH
10kHz	1·6μF	0·16mH
15kHz	1μF	0·11mH
20kHz	0·8μF	0·08mH

PRINCIPLE OF OPERATION

The design of the Universal Bridge is based on the transformer ratio-arm principle. A full explanation of the theory of operation is given in Wayne Kerr Monograph No. 1, 'The Transformer Ratio-Arm Bridge', available on request.

Balance of the Unknown impedance is against Standards of conductance and capacitance in parallel. Tappings on the two Bridge transformers, connected to decade controls, permit measurements to be made accurately on a wide range of impedance in any quadrant of the complex plane.

An internal oscillator adjusted to 1592 Hz ($\omega = 10^4$) provides the Source voltage. A buffer amplifier isolates the oscillator from the Bridge circuits, to which four-terminal connexions can be made. The Detector is a tuned two-stage amplifier, incorporating a sensitivity control, with a double-shadow 'magic eye' associated with each stage. Instruments can be supplied to special order with Source and Detector tuned to frequencies other than 1592 Hz.

A simplified diagram of the circuit arrangement is shown in Fig. 12, where Z_u and Z_s are the Unknown and Standard impedances respectively.



Fig. 12. Principle of Operation

Balance will be indicated by the Detector, connected to the secondary of the current transformer $T2$, when equal currents flow from either end of the centre-tapped primary of $T2$. In this condition, the potential across the primary will be zero and, therefore, the right-hand terminals of Z_u and Z_s will be at Neutral potential. Thus the same voltage is applied to both impedances, and for equal currents to flow in the two halves of T2 primary, the resistive and reactive components of the Unknown impedance must be equal to those of the Standard.

For a given value of Unknown impedance the Bridge can be balanced in three ways. The value of the Standards can be adjusted, differing potentials can be applied to the two impedances, or the currents they pass can be fed through unequal numbers of turns on the primary of $T2$. A selection of taps on $T2$ primary and on the secondary of the voltage transformer $T1$ permit a wide range of Unknown values to be measured against a minimum number of Standard components.

The Standard components are resistive and capacitive but, by reversing the sense of their connexions to the current transformer, an impedance in any quadrant of the complex plane can be measured. As the Standards are connected in parallel, the Bridge measures the equivalent parallel components of the Unknown.

Since at balance, no potential exists across $T2$ primary, it is possible to connect an impedance between the right-hand terminal of Z_u and Neutral without affecting the accuracy of measurement. The only effect will be to reduce the sensitivity of the Detector and this can be compensated for by increasing the gain.

It is possible also to connect an impedance between the left-hand terminal of Z_u and Neutral. This shunt impedance will reduce the voltage applied to Z_u, but will reduce also the voltage applied to Z_s, in proportion to the turns ratio. The accuracy is therefore unaffected and any loss in sensitivity can again be compensated for by increasing the detector gain.

The transformers are so designed that very heavy shunting is possible without seriously affecting the accuracy of measurement. Components can in most instances, therefore, be measured *in situ*.

CIRCUIT DESCRIPTION

POWER SUPPLY

The Instrument operates from supplies of 50 to 60 Hz, consuming about 25 watts. A circuit diagram is provided at the back of this Handbook and from this it can be seen that the power transformer is tapped to suit supplies of 100–125 and 200–250V. The full-wave rectifier, V4, supplies approximately 15mA at 250V.

SOURCE

Refer to Fig. 13. One section of the double triode operates in an LC oscillator circuit and a fraction of the alternating potential is fed at low impedance to the right-hand section of *V*1, serving as a buffer amplifier. Inductor *L*2 has a ferrite pot core and the frequency of oscillation is set to precisely 1592 Hz by adjustment of this core. A potential of approximately 30V is developed across *T*1 secondary.



Fig. 13. Source Circuit

DETECTOR

The Detector circuit, shown in Fig. 14, consists of a two-stage tuned-anode amplifier using ferrite pot cores and silvered-mica capacitors. A double-shadow 'magic eye' is connected to each stage, providing four degrees of sensitivity overall. The Sensitivity control (*RV*4) affects both 'magic eyes' and balance is indicated by maximum shadow.



Fig. 14. Detector Circuit

BRIDGE CIRCUIT

The Bridge circuit, shown in simplified form in Fig. 15, consists essentially of the voltage transformer *T*1, tapped to provide an accurately-related selection of voltages across the Unknown and Standard impedances, and the current transformer *T*2.

The primary of *T*2 consists of two separate windings on a core of high-permeability metal. One winding is fed from the Unknown impedance and the other from the Standards circuit. Balance occurs when the ampere-turns in the two windings are of equal magnitude but opposite sense.

The Neutral connexion from the voltage transformer is taken to the centre-tap of the Standards winding on the current transformer. By connecting the Standards to a winding of appropriate sense, positive or negative conductance and capacitance can be measured.

The Neutral connexion of the Unknown winding on the current transformer is linked to that of the Standards winding for most two- and three-terminal measurements. The link is removed for four-terminal measurements, to provide two Neutrals.



Fig. 15. Bridge Circuit

STANDARDS

For details of the Standards arrangement the complete circuit diagram at the back of this Handbook should be referred to.

Switched decades are provided by two fixed resistors (*R*15, *R*16) and two fixed capacitors (*C*12, *C*14) which can be switched independently in 1-turn steps to taps of from 0 to 10 turns on the voltage transformer. This gives each Standard effectively ten different values.

To provide the continuously-variable conductance decade, a linear potentiometer (*RV*2) is connected across nine turns of the voltage transformer and produces a variable voltage across a close-tolerance resistor (*R*17). This is illustrated in Fig. 16. The continuously-variable reactive decade is provided by a variable air-spaced capacitor (*C*11) connected to the 3-turn tap on *T*1.

The values of the conductance and capacitance Standards are so chosen that they normally cover the same range. However, to extend the effective range of one Standard with respect to the other, *T*2 is tapped to provide a capacitance range one-tenth of normal. The range multiplier switches have the following settings:

G 1 and —1. (—1 for negative conductance).

C 0·1, 1 and —1. (—1 for inductance).



Fig. 16. Third *G* Decade

TRIMMING OF STANDARDS

The two main resistive Standards (*R*15, *R*16) are wire-wound and slightly inductive. Their phase angle is reduced to zero by means of trimming capacitors (*C*36, *C*15 and *C*16) connected in parallel with them.

The 0·001μF Standard (*C*12) does not require correction. Losses in the 0·01μF Standard (*C*14) are compensated for by a resistive current fed into the current transformer through a tap of opposite sense. This is shown in Fig. 17. *RV*1 is adjusted to balance out the resistive current produced by the losses in *C*14. *R*12, *R*13 and *R*14 form an attenuator to obviate the use of high-value resistors which might become unstable. *R*13 is always connected to a tap having the same number of turns as that connected to *C*14 but of opposite sense. The trimming holds, therefore, for all positions of the *C* multiplier switch.

of a fixed 33μμF capacitor (*C*18) and a 100μμF variable capacitor (*C*17) connected across the positive and negative ends of the Standards winding on the current transformer. The same voltage as that used for conductance trimming is fed to their junction. With this arrangement the effective value of *C*17 can be made either positive or negative.



Fig. 17. *C* Phase Trimming

RESIDUAL CAPACITANCE TRIMMING

To compensate for the residual capacity of the continuously-variable Standard *C*11, a trimmer (*C*13) feeds a current into *T*2 through a tap of opposite sense. The arrangement is shown in Fig. 18. Whatever the connexion of *C*11 to the current transformer, the trimmer *C*13 is always connected to a tap having the same number of turns but of opposite sense. Once *C*13 is adjusted, therefore, the trimming holds for any position of the *C* multiplier switch.



Fig. 18. Residual Capacitance

MAIN BRIDGE TRIMMING

The circuit arrangement of the two external trimming controls (Trim *G*, Trim *C*) is shown in Fig. 19.

A compensating voltage from the 9-turn tap on *T*1 is fed via *R*18 to the slider of a potentiometer (Trim *G*) connected across the +100 and —100 taps of *T*2. This allows the effective value of *R*18 to be made either positive or negative.

The capacitive trimming consists



Fig. 19. Trim Controls

UNKNOWN CONNEXIONS AND RANGE SWITCH

The circuit arrangement of the connexions to the Unknown impedance and the facilities for range changing are shown in Fig. 20.

The impedance to be measured is connected between the voltage transformer and the Unknown winding of the current transformer via the Range switch (*S*7). The Range switch connexions to the taps on the transformers are such that in each of the seven positions the preceding range is effectively multiplied by ten.

Physical connexion to the Unknown is by means of two external coaxial cables. The screen of each cable is connected to the associated Neutral of the Bridge. For two- and three-terminal measurements the two Neutrals are linked: this method of connexion prevents the inherent capacitance of the cables shunting the measurement terminals.



Fig. 20. Range Switch

MEASUREMENT VOLTAGE

The a.c. voltage applied to the component being measured on the Bridge cannot be specified accurately, measurements being made by comparative values of the Standard and Unknown currents. The voltage varies from range to range but is constant over any one range. Approximate values are:

Bridge Range	R.M.S. Voltage
1	30 V
2 and 3	3 V
4 and 5	0·3 V
6 and 7	30mV

MAINTENANCE AND SERVICE

Important: *No attempt should be made by unskilled personnel to service the Bridge. Maintenance, and particularly any trimming using the pre-set controls, must be carried out by a qualified service engineer in accordance with the following instructions.*

SOURCE AND DETECTOR

The location of any fault in the oscillator or detector amplifier circuits will be indicated by checking the valve potentials against the values shown in Table 5. These readings were obtained with an AVOmeter model 7 (negative lead to chassis) and are average values: variations of up to ±20 per cent must be expected.

TABLE 5

Valve	Electrode	Pin No.	Avo Range	Voltage
V1	Cathode 1	8	100V d.c.	35V
V1	Cathode 2	3	10V d.c.	4·2V
V1	Anode 1	6	400V d.c.	250V
V1	Anode 2	1	400V d.c.	120V
V2	Cathode	2	10V d.c.	1·5V
V3	Cathode	2	10V d.c.	1·1V
V3	Screen	7	400V d.c.	75V
V4	Cathode	7	400V d.c.	255V

OVERALL GAIN

To check the overall gain, first set-up the Bridge according to the procedure given on page 10 but with the Range switch on position 3. Using two 100-ohm (20 per cent) resistors, connect one between inner and neutral of the *E* lead and the other across the *I* lead. Leave the Sensitivity control at maximum and adjust the variable conductance control to read 5(0·05μ Mho). The most sensitive 'magic eye' section should close completely. This degree of sensitivity is required only when the Low Impedance Adaptor is in use: for all normal Bridge measurements one-tenth of such sensitivity provides adequate discrimination.

OSCILLATOR FREQUENCY

The oscillator frequency is adjusted in the final test procedure to 1592 Hz ±0·25 per cent. If a Frequency Standard or Frequency Counter is available, it can be fed with the measuring potential from the Bridge. The Source frequency will be obtained with an accuracy determined by the Standard or Counter.

An alternative method, depending only on the Bridge itself and two components, can be used if the Bridge Standards have been checked (see pages 25 and 26). The components required are a high-stability resistor and a high-precision mica capacitor. The exact values are not important but should be about 10$k\Omega$ and 0·01μF (1$k\Omega$ and 0·1μF could be used as an alternative). The method is as follows:

1 Measure the resistor and record the value of conductance G.

2 Measure the capacitor and record the value of capacitance C.

3 Connect the two components in series and measure the effective conductance G_m and effective capacitance C_m of the combination.

It can be shown that

$$\omega = \left[G_m/[(C^2/G)\,(1-G_m/G)]\right]^{\frac{1}{2}} \quad . \quad . \quad .(1)$$

or $$\omega = (G/C\,.\,[(C/C_m)-1]^{\frac{1}{2}} \quad . \quad . \quad . \quad . \quad .(2)$$

Equation (1) or (2) can be used to establish accurately the angular frequency of the Bridge Source. The tables used for the calculation should preferably be five-figure. If the Bridge Standards were checked correctly, the values of G. C, G_m and C_m should be obtained with an accuracy of ±0·1 per cent. The angular frequency will then be obtained with an accuracy of better than ±0·3 per cent. If necessary, re-set the core of L.2.

BRIDGE ACCURACY

An important feature of the Universal Bridge is the ease with which the accuracy can be checked and, if necessary, the Standards re-set. As described earlier, the B221A uses only one, fixed, Standard per decade; the steps of the decades, and the Bridge ratios, being obtained from the ratio-transformers. The Standards fitted to the Bridge are adjusted to an accuracy of 0·05 per cent and should have a long-term stability of the same order. The ratio-transformers have an accuracy an order higher, about 0·01 per cent.

Before the Standards are checked, the following points should be noted. If high accuracy is required in measuring any component, all three decade controls must be in use. There is no requirement for these three controls to be checked to the same order of accuracy. Ignoring the decimal point and units, consider the reading 1234. The 1 is given by the first decade, the 2 by the second, and the figures 34 by the continuously-

variable control. The reading accuracy is within 0·1 per cent. If the Standard of the first decade is 0·1 per cent in error, the true value could be 1235 or 1233. The same effect would be caused by an error of 0·5 per cent in the second decade, or 3 per cent in the continuously-variable control.

The highest reading accuracy is obtained when all three decades are at maximum, giving a reading of 11110. With the Sensitivity control at maximum, a 1 per cent change in the continuous variable can be noticed. This corresponds to 0·1 per cent in the second decade and 0·01 per cent in the first decade. The highest accuracy is obtained, therefore, by adjusting the Standard in the first decade to 0·01 per cent, the Standard in the second to 0·1 per cent, and the continuously-variable control to 1 per cent. If this is done, the Bridge can often be used for measurements to an accuracy of 0·02 per cent.

One further point to be noted is that an aged Standard has a better long-term accuracy than a new one. Consequently, whenever possible the existing Standards should be adjusted by means of series or shunt elements in preference to being replaced.

CHECKING AND ADJUSTING STANDARDS

The absolute accuracy of the Bridge can be determined using only two external fixed Standards, one a resistor, the other a capacitor. As the Bridge Standards are themselves adjusted to 0·05 per cent, the reference Standards must be known with certainty to at least this degree of accuracy. The ideal reference Standards are a 0·01 per cent resistor of either 1000 or 10 000 ohms and a 0·01 per cent capacitor of either 0·1 μF or 0·01 μF. Using a reference of less than 1000 ohms (resistance or reactance) causes difficulty with the series impedance of the leads etc.; above 10 000 ohms trouble can arise from the spurious shunt elements. Even with a 0·01 μF Standard the connexions to the Bridge must be made carefully because 1 $\mu\mu F$ represents an error of 0·01 per cent, and this capacitance is produced by two inches of wire exposed outside the Neutral screen.

CONDUCTANCE STANDARDS

The procedure for checking the conductance Standards of the Bridge is described for a 1000-ohm external reference. If a 10 000-ohm reference is used, the Bridge range settings stated must be reduced by 1.

1 With the measurement leads (Neutrals linked) in position, and the *G* and *C* multipliers both set to 1, switch to Range 6, set all capacitance controls to zero, the two switched conductance decades to zero, but set the continuously-variable conductance control at 06 (·006μMho).

2 Adjust the two Trim controls for exact balance with the Sensitivity control set fully clockwise. The reason for setting the variable control at 06 is to ensure that, when the Standard used in the first decade is checked, balance is possible without using the un-checked second decade, whether the Standard be slightly high or low.

3 Connect the $1k\Omega$ reference resistor to the measurement leads and balance the Bridge by means of the *first* and *third* conductance controls only. To obtain a clean balance it may be necessary to adjust the Trim *C* control or to change the *C* multiplier switch setting to —1 and, after any necessary re-trimming, to operate the variable *C* control. If the exact point of balance lies between 1·005 and 1·007 mMho, the first decade Standard is within the specified tolerance and no adjustment is necessary.

If balance does not occur within these limits, suitable series or shunt resistors should be added to the Standard in the Bridge. These resistors do not have to be close tolerance.

If it is desired to adjust the first decade Standard to 0·01 per cent, carry out the procedure described in 3a and 3b.

3a Disconnect the reference resistor and repeat steps 1 and 2 with the Range switch in position 5.

3b Re-connect the reference resistor, set the first conductance decade to 10 (⊕) and, if necessary, add series or shunt resistors to the first decade Standard until balance occurs between 1000·5 and 1000·7μMho. Adjust the *C* controls as described in para. 3.

4 Disconnect the reference resistor and repeat step 1.

5 Adjust the two Trim controls for exact balance with the Sensitivity control set fully clockwise.

6 Re-connect the reference resistor, set the second conductance decade to 10 (⊕) and balance the Bridge using the continuously-variable control. The exact point of balance should be between 1·005 and 1·007mMho. If necessary, add series or shunt resistors to the Standard of the second decade to bring the balance point within this range.

7 Set *all* decade controls at zero, disconnect the reference resistor and switch to Range 7. With the Sensitivity control fully clockwise, adjust the two Trim controls for a precise balance.

8 Re-connect the reference resistor and balance the Bridge. The reading should lie within half a scale division of the 10 calibration mark (1mMho). If outside this tolerance, set to 10 and balance the Bridge by adjusting the pre-set potentiometer *RV*5.

The calibration points 1 to 9 on the continuously-variable control can be checked quickly with a decade resistor box but this item is not essential. The alternative procedure is as follows.

9 Zero the Bridge on Range 3 and then set the first decade control to 1 (1·000μMho).

10 Connect the outer terminals of a 10*k*Ω potentiometer to the inner and Neutral respectively of the *E* measurement lead. Connect one end of a 100*k*Ω (20 per cent) resistor to the inner of the *I* measurement lead and the other end of this resistor to the wiper terminal of the potentiometer.

11 With the Sensitivity control set at mid-position, adjust the potentiometer to balance the Bridge.

12 Without disturbing the potentiometer adjusted in para. 11, change to Range 5, set the first decade to zero, and balance the Bridge using the continuously-variable control. Balance should occur within half a scale division of the 1 calibration mark.

Repeat the procedure described in paras. 9, 10, 11 and 12 with settings at 2, 3, etc. to 9. If desired, intermediate points can be checked by using both switched decades in para. 9.

By means of successive adjustments to the pre-set potentiometer *RV*5, it should be possible to bring the accuracy at all settings to within half a scale division. Any serious departure from this calibration is most likely to be due to a worn track on potentiometer *RV*2, in which instance it should be replaced. Play in the drive mechanism linking the potentiometer to the dial can also cause trouble. The degree of engagement between the racks and pinions can be adjusted by repositioning an eccentric cam.

CAPACITANCE STANDARDS

Succeeding paragraphs describe the procedure for checking the capacitance Standards of the Bridge, using a 0·1 μ*F* external reference. If a 0·01μ*F* reference is used, the Bridge range settings stated must be reduced by 1.

1 With the measurement leads (Neutrals linked) in position, and the *G* and *C* multiplier switches both set to 1, switch to Range 6, set all conductance controls to zero, the two switched capacitance decades to zero, but set the continuously-variable capacitance control at 2(·002μ*F*).

2 Adjust the two Trim controls for exact balance with the Sensitivity control set fully clockwise. The reason for setting the variable control at 2 is to ensure that, when the Standard used in the first decade is checked, balance is possible without using the unchecked second decade, whether the Standard be slightly high or low.

3 Connect the 0·1μ*F* reference capacitor to the measurement leads and balance the Bridge by means of the *first* and *third* capacitance controls only. To obtain a clean balance it may be necessary to adjust the Trim *G* control. If the exact point of balance lies between ·1019 and ·1021μ*F*, the first decade Standard is within the specified tolerance and no adjustment is necessary.

If balance does not occur within these limits, remove *C*37 and replace it with a low-loss capacitor of suitable value.

If it is desired to check the first decade standard to 0·01 per cent, carry out the procedure described in 3a and 3b.

3a Disconnect the reference capacitor and repeat steps 1 and 2 with the Range switch in position 5.

3b Re-connect the reference capacitor, set the first capacitance decade to 10 (⊕) and, if necessary, make further changes to the value of *C*37 until balance occurs between ·10019 and ·10021μ*F*. Adjust Trim *G* if necessary.

4 Disconnect the reference capacitor and repeat step 1.

5 Adjust the two Trim controls for exact balance with the Sensitivity control set fully clockwise.

6 Re-connect the reference capacitor, set the second capacitance decade to 10(⊕) and balance the Bridge using the continuously-variable control and Trim *G*. The exact point of balance should be between ·1019 and ·1021μ*F*. If outside these limits, unseal trimmer *C*38 and adjust this component as necessary.

7 Set *all* decade controls at zero, disconnect the reference capacitor, and switch to Range 7. With the Sensitivity control fully clockwise, adjust the two Trim controls for a precise balance.

8 Re-connect the reference capacitor and balance the Bridge using the continuously-variable capacitance decade. The reading should lie within half a scale division of the 10 calibration mark (·1μ*F*). If outside this tolerance, first check that the racks and pinions are engaging properly and if necessary adjust the eccentric cam. The relative settings of the dial and variable capacitor can be altered slightly but, if this is done, *C*13 will have to be re-set as described in step 10.

9 Set *all* decade controls at zero, disconnect the reference capacitor, switch to Range 1 and, with the Sensitivity control set fully clockwise, adjust the two Trim controls for a precise balance.

10 Change the *C* multiplier switch position from 0 to 0·1 and to —1. The Bridge should remain balanced on all three positions. If it does not, unseal trimmer *C*13 and make a small adjustment. Repeat the process and adjust Trim *C* until the setting is found where balance is maintained on all three switch positions.

The calibration points 1 to 9 on the continuously-variable control can be checked quickly with a decade capacitor box but this item is not essential. The alternative procedure is as follows:

11 Zero the Bridge on Range 2 and then set the first capacitance decade control to 1 (10μμ*F*). Connect a variable capacitor, covering the range 10–90μμ*F*, to the measurement leads.

12 With the Sensitivity control set at mid-position, adjust the variable capacitor to balance the Bridge.

13 Without disturbing the variable capacitor, change to Range 4, set the first decade to zero and balance the Bridge using the continuously-variable control. Balance should occur within half a scale division of the 1 calibration mark.

Repeat the procedure described in paras. 11, 12 and 13 with settings at 2, 3, etc. to 9. If desired, intermediate points can be checked by using both switched decades in para. 11.

Should the calibration error exceed half a division at some points, adjustment should be made to the mechanism linking *C*11 to the dial. This adjustment *must* be followed by a repetition of the procedure described in paras. 9 and 10.

PHASE CORRECTION

Succeeding paragraphs describe firstly the adjustment of the two trimmer capacitors which off-set the slightly inductive properties of the two wire-wound conductance Standards and, secondly, the adjustment of the pre-set potentiometer which controls the current compensating for the losses associated with the larger capacitance Standard.

1 Zero the Bridge on Range 4.

2 Select a 10*k*Ω *carbon* resistor slightly below the nominal value, so that balance is obtained with the first and last *G* decades only (the centre decade *must* remain at zero).

3 Adjust trimmer *C*16 to obtain a clean balance.

4 Change to Range 5, return the first decade to zero and set the second decade at 10 (⊕).

5 By means of the variable *G* decade and trimmer *C*36, obtain a clean balance.

The pre-set potentiometer *RV*1 is adjusted in the Test Department while a 3-terminal 10μμ*F* absolute Standard is connected to the Bridge. The potentiometer is not likely to need re-adjustment and, unless a capacitor of accurately-known power factor is available, the original potentiometer setting should not be disturbed.

If the external reference capacitor used for checking the decades has a tolerance of 0·01 per cent, the power factor will probably be quoted (at 1000 Hz). This factor will be virtually unchanged at 1592 Hz. The capacitor could be used, therefore, to check the setting of *RV*1.

Part 2 - Low Impedance Adaptor Q221A

INTRODUCTION

The lower limit of impedance measurement that can be made accurately with the Universal Bridge is determined by lead resistance and the winding resistance of the Bridge transformers. The Low Impedance Adaptor, designed to operate in conjunction with the Bridge, extends the range of measurement to include large values of capacitance and very small values of resistance and inductance.

Measurements with the Bridge itself are of the equivalent *parallel* components of conductance and capacitance, obtained by applying a reference potential to the Unknown and measuring the resulting current. The function of the Adaptor is to reverse the procedure to one of passing a reference current through the Unknown and measuring the resulting voltage. This reversal is achieved with a resistive *T*-network where the shunt arm consists of the Unknown. The required value is derived very simply from the Bridge dial-readings in terms of the equivalent *series* components of resistance and inductance.

The prime function of the Adaptor is to permit measurements on impedances of less than ten ohms to be made *with the first decade switches set permanently to zero*. Under these conditions the value of the Unknown has a negligible effect on the reference current and measurements are of the specified accuracy. Only when absolute accuracy is a secondary consideration to the detection of small changes should the first decade be used. The usefulness of very fine discrimination and measurements from 10 to 100 ohms are described in the text.



SPECIFICATION

Measurement Ranges

Range	R		L‡		C§	
	First Division	Maximum*	First Division	Maximum*	Minimum*	First Division
1	20 μΩ	10 mΩ	·002 μH	1 μH	10,000 μF	5F
2	200 μΩ	100 mΩ	·02 μH	10 μH	1,000 μF	·5F
3	2 mΩ	1 Ω	·2 μH	100 μH	100 μF	·05F
4	20 mΩ	10 Ω	2 μH	1 mH	10 μF	·005F

Accuracy

R: ±1% ±25 μΩ.
L: ±1% ±0·005 μH.
C: Frequency-dependent (see Text).

* Values quoted apply with first G and C decade switches at zero (see Text). With all three decades in use: R maximum and L maximum are increased by a factor of 10, and C minimum is decreased by a factor of 10.

‡ All L values decreased by a factor of 10 with multiplier on C×0·1. (Not available from AA221.)

§ All C values apply only at 1592 Hz. See text for other frequencies.

Discrimination	R:	0·2% of Maximum (see Table).	
	L:	0·2% of Maximum (see Table).	
	C:	Frequency-dependent (see Text).	
Dimensions	Base Diameter:	$5\frac{1}{16}$ in.	(12·9 cm.).
	Height:	$3\frac{5}{16}$ in.	(8·5 cm.).
Packed Weight	Not exceeding	$5\frac{1}{2}$ lb.	(2·5 kg.).

OPERATING INSTRUCTIONS

MEASUREMENT PROCEDURE

1 Remove the measurement cables from the Bridge and connect them to the sockets provided on the Low Impedance Adaptor.

2 Connect the two cables from the Adaptor to the corresponding sockets on the Bridge.

3 Turn the Bridge Range switch to position 3. The Bridge must *always* be on this Range when the Low Impedance Adaptor is in use.

4 Turn the Low Impedance Adaptor to the desired range and connect the two leads from the 'blue' socket to either side of the impedance to be measured.

5 Connect the neutral (green) lead from the 'white' cable between the unknown impedance and the neutral clip from the 'blue' cable.

6 Clip the 'inner' (red) lead from the 'white' cable on to the neutral clip of this same cable. This provides the necessary condition of only a single-point contact between one side of the Unknown and the detector for the initial trimming adjustment.

7 Set the Bridge decade controls to zero and adjust the Trim G and Trim C controls for maximum shadow on the magic eyes.

8 Transfer the red lead from the 'white' cable to the other side of the Unknown. The points of connection for the two leads from the 'white' cable must be immediately adjacent to the impedance it is desired to measure. The two connections from the 'blue' cable should lie outside these. This is illustrated in Fig. 21 (below right) where the reference current is obtained from the 'blue' cable and connection to the 'circuit measuring voltage' is by the 'white' cable.

9 Set the G and C switches as follows:
both at 1 for R and L,
G at 1 and C at –1 for R and C,
either or both at –1 for mutual or transfer impedance.

10 Adjust the second switched decades and vernier controls of the Bridge to obtain maximum shadow on the magic eyes. The first switched decades should not be employed if the measured value is to be derived from the simplified expressions shown for each of the four ranges on the Low Impedance Adaptor.



Fig. 21. Connexion to the Unknown

INTERPRETATION OF RESULTS

The equivalent series components of the unknown impedance are derived from the Bridge dial readings at balance by using the conversion formulae shown for each of the four ranges on the Low Impedance Adaptor. These formulae are reproduced in Fig. 22.

Capacitance can be also be derived from the Conversion Chart on page 35. In either case, however, the results apply only at 1592 Hz (see '50 Hz—20 kHz Operation' on page 32).



Fig. 22. Adaptor Range Plate

Example (R) Adaptor on 0–1Ω range
(as shown on Q221A dial)
Bridge reading 0·337μMho
Adaptor range plate:
$\Omega = \pm G/10$
$= 0{\cdot}337/10$
or $R_u = 0{\cdot}0337\Omega$

Example (L) Adaptor on 0–10000μH range
(as shown on Q221A dial)
Bridge reading 64·8μμF
Adaptor range plate:
$\mu H = +C \times 10$
$= 64{\cdot}8 \times 10$
or $L_u = 648\mu H$

Example (C) Adaptor on 1–100μF range
(as shown on Q221A dial)
Bridge reading –17·4μμF
Adaptor range plate:
$\mu F = 10^3/-C$
$= 1000/-(-17{\cdot}4)$
or $C_u = 57{\cdot}5\mu F$

MEASUREMENT OF REACTANCE

The Bridge is so designed that the resistive and reactive scales cover the same range in ohms. If on a particular range of the Adaptor the Bridge *G* dials cover 0 to 1 ohm, the *C* dials, therefore, cover 0 to 1 ohm also. Assuming the Bridge has been balanced, the procedure for measuring reactance directly in ohms is as follows.

Mentally transfer the figures on the C dials to the *G* dials (so that the last digits are coincident) and insert in this transferred reading a decimal point where one occurs on the *G* dials. Apply to this amended reading the scale multiplying factor for ohms given on the Adaptor range plate. The result is the required reactance value.

Example:
Adaptor range plate set to 0–100Ω.
Bridge reading 0·362μM*ho* and 117·4μμ*F*.
If the figures that appear on the *C* dials (giving the reading 117·4 μμ*F*) had appeared on the *G* dials, the reading would have been 1·174μ*Mho*. This is seen by writing them underneath in corresponding positions:

0·362μ*Mho* (actual reading on *G* scale)
1·174μ*Mho* (figures mentally transferred from *C* scale.)

The range plate on the Adaptor gives
$\Omega = \pm G \times 10$
giving for the resistance term
$R_u = 0{\cdot}362 \times 10$
$= 3{\cdot}62$ ohms
and for the reactance term
$X_u = 1{\cdot}174 \times 10$
$= 11{\cdot}74$ ohms

CONNEXION TO THE UNKNOWN

When very low values of impedance are to be measured *in situ*, some precautions must be observed or large errors may arise. Two faults are indicated by the left-hand illustration of Fig. 21. Firstly, if the two cables are arranged side-by-side and large loops formed at their ends, there is considerable danger that the reference current in one loop may induce a sufficiently large voltage in the other to cause a significant error. Secondly, the voltage being measured is not only the potential drop across Z_u but also that occurring in the leads from Z_u to the point of contact of the current clips.

In this respect it must again be emphasized that the lead coded blue, from the Adaptor to the Unknown, although associated with the *E* plug of

the Bridge is effectively the CURRENT lead from the Adaptor. Similarly, the function of the white (*I*) lead is to feed back to the Bridge the VOLTAGE measured across the Unknown.

The areas of the loops involved in the connexions should be kept as small as possible and the coupling between them made as low as possible. This is shown diagrammatically in the right-hand illustration of Fig. 21, which shows also the correct positioning of the current clips outside the voltage clips. These connexions must be made independently on to the leads from the Unknown: the clips must not touch and one set must never be used as a connexion for the other.

ERRORS IN THE MINOR COMPONENT

It must be realised that inaccuracies may occur when measuring the minor component of a complex impedance whose resistive and reactive elements are very different in magnitude at 1592 Hz. (It may be noted that a minor component of 0·1 per cent affects the modulus of the impedance by less than one part in a million.) The Adaptor is not suitable for accurate measurements on small minor components, but a reliable approximation can be obtained if great care is taken in the adjustment of the Trim controls and the leads are disposed for minimum coupling and for minimum disturbance when changed from the trim to the measurement connexions.

A correction to allow for the residual reactance of the Adaptor can be made as follows. The measured values of the resistance and reactance of the Unknown are first obtained in the usual manner. The Unknown is then replaced by a carbon resistor of near value to the resistance of the Unknown, using the shortest possible connexions from the terminals or clips. The resistance and apparent reactance of this resistor are measured. As a carbon resistor has a totally insignificant reactive term at the Bridge frequency, the reactance must be due to the Adaptor or Bridge circuits. In both measurements a note must be made of the sign of the reactance (positive for inductive and negative for capacitive). The true reactance of the Unknown is given by:

(Apparent reactance of Unknown)—(Apparent reactance of carbon resistor).

This true value of reactance can be converted into the corresponding inductance or capacitance from the appropriate expression:

$$L = (\text{True reactance})/10^4,$$

$$\text{or } C = 1/(10^4 \times \text{True reactance}).$$

MEASUREMENT OF R.F. COILS

It must be borne in mind that, although measurements can be made in the millimicrohenry region, the Adaptor is not ideally suited to the measurement of r.f. and v.h.f. coils. This is because skin effect and self-capacitance result in the inductance at the operating frequency of the coils being very different from the value at 1592 Hz. The change is dependent upon so many factors (geometry, wire size, spacing, type of insulation, effect of screening and so on) that no useful extrapolation can be made. Furthermore, air-cored coils of small inductance behave, at audio frequencies, like slightly inductive resistors, so that accurate measurement of a minor component would be required.

The Adaptor is useful for comparison measurements and for obtaining the coupling coefficients of i.f. transformers, sections of delay line, etc. Also, in v.h.f. circuits a large proportion of the circuit inductance is contained, often, in the wiring and switches. The geometry in such instances makes possible reasonably accurate measurements using the Adaptor and, in fact, the value of such inductance is difficult to ascertain otherwise because the measurement must be made *in situ*.

MEASUREMENT OF MUTUAL INDUCTANCE

The method of connecting a transformer is illustrated in Fig. 23, where the two resistors are the Standard components in the Adaptor. Assuming that the loading imposed by the secondary *R* is insignificant and that the self-impedance of the primary is so low that it has no effect on the current, the current through the primary will be



Fig. 23. Measurement of Mutual Inductance

given by E/R. This will induce into the secondary a voltage of $\pm j\omega M.E/R$, where the sign is determined by the relative sense of the windings. At balance the induced voltage will drive a current, $\pm j\omega M.E/R^2$, through the current transformer of the Bridge. Form equation (1), under Principle of Operation, (page 34):

$$\pm j\omega M = R^2.I/E.$$

Replacing I/E by the equivalent on the Standards side of the Bridge:

$$j\omega M = \pm j\omega C \,.\, R^2$$

$$\text{or } M = \pm C \,.\, R^2$$

where R^2 is 10^4, 10^5, 10^6 and 10^7 ohms for Q221A Ranges 1-4, respectively.

Mutual inductance is read, therefore, in the same manner as self-inductance, the alternative signs being catered for by the 1 and —1 positions of the C selector switch. The setting required depends on the relative sense of the two windings.

POLARIZING ELECTROLYTIC CAPACITORS

When it is considered necessary to prevent a reverse polarity (caused by the small a.f. measuring



Fig. 24. Polarizing Electrolytic Capacitors

voltage) from appearing across an electrolytic capacitor, a polarizing potential of one or two volts from for example a single dry cell can be applied as shown in Fig. 24. Under no circumstances should a high polarizing voltage be used.

ACCUMULATOR RESISTANCE

The Low Impedance Adaptor can be used for certain measurements on accumulators. A limitation is imposed by the permissible d.c. through the Adaptor resistors. These have a continuous rating of 0·5 watt. The maximum values of d.c. voltage that may be applied to the *Adaptor* terminals (these values *must not* be applied directly to the Bridge), and the corresponding current values, are given below.

Adaptor Range	Maximum d.c. Voltage	Maximum Direct Current into Adaptor
1	7V	140 mA
2	12V	80 mA
3	22V	45 mA
4	40V	25 mA

It must be realised that the measurement will be of the impedance of the accumulator, consisting of a resistance in series with a small reactive component.

50Hz - 20kHz OPERATION

Resistances of less than 10 ohms are measured in the normal manner.

Inductances of less than $1mH$ (or $10mH$ if the first decade control is operated) are measured on the Low Impedance Adaptor. When the value exceeds $1mH$ (or $10mH$) the measurement can be made with the Bridge alone provided that a sufficiently high frequency is employed (see table 4, page 18).

Capacitances whose reactance falls below 10 ohms at any one frequency (the dividing line is indicated in table 4) are measured with the Low Impedance Adaptor but the relationships for capacitance, given on the Adaptor range plate, do *NOT* apply at any frequency except 1592 Hz. The value of the unknown capacitance is given by the expression

$$C_u = -1/(\omega^2 R^2 C_b)$$

where R^2 is 10^4, 10^5, 10^6 and 10^7 ohms for Q221A Ranges 1-4, respectively.

At a frequency of 1592 Hz ($\omega = 10^4$) the above expression can be simplified to give the scale multiplying factors for capacitance shown on the Adaptor range plate. These same four factors can be used for computing the measured capacitance at other frequencies if the result is multiplied by a suitable constant (K). This is equivalent to re-writing the above expression as:

$$C_u = K[-1/(\omega_o^2 R^2 C_b)]$$

where $\omega_o = 2\pi \times 1592$ and $K = (\omega_o/\omega)^2$.

The value of K at selected frequencies between 50 Hz and 20kHz is shown in the table (below).

Measurement Frequency	K	Measurement Frequency	K
50Hz	1015	1592Hz	1·00
60Hz	704	2kHz	0·634
100Hz	254	4kHz	0·158
200Hz	63·4	6kHz	0·0704
400Hz	15·8	8kHz	0·0397
600Hz	7·04	10kHz	0·0254
800Hz	3·97	12kHz	0·0176
1000Hz	2·54	15·92kHz	0·01
1200Hz	1·76	20kHz	0·0063

Example

With the range multiplier switches set at $1(G)$ and -1 (C), balance was obtained on Range 3 of the Adaptor with a capacitance dial reading of $-74{\cdot}1\mu\mu F$.

The measurement frequency was 800 Hz.

From the Adaptor range plate:

$$\mu F = 10^4/-C \text{ (at 1592 Hz)}$$
$$= K[10^4/-C] \text{ at 800 Hz.}$$

From the table, K at 800 Hz is 3·97.

$$\therefore C_u = 3{\cdot}97\ [10^4/-(-74{\cdot}1)]\mu F$$
$$= 3{\cdot}97 \times 135\mu F$$
$$= 536\mu F.$$

With a Bridge dial reading of $10\mu\mu F$, the maximum values of capacitance that can be measured at a frequency of 400 Hz are:

Range plate at 1–100$k\mu F$	1·58 Farads
Range plate at ·1–10$k\mu F$	158 000μF
Range plate at 10–1000μF	15 800μF
Range plate at 1–100μF	1580μF

The range of capacitance measurement is further extended by the use of lower measurement frequencies. Also, if the Bridge dial reading at balance is less than $10\mu\mu F$, the measured capacitance is of a very high value. Such measurements are obtained with reduced sensitivity and accuracy.

When very large values of capacitance are measured, the inductance of connecting leads becomes of extreme importance and must be kept to the absolute minimum. For example if a capacitor of 200 000μF is measured at 1 kHz, a lead inductance of only $\frac{1}{8}\mu H$ will resonate with the capacitor. Even $^1/_{80}\mu H$ will cause a 10 per cent error in the measured capacitance value.

Note: *ALL masuerements with the Low Impedance Adaptor* must *be made with the Bridge on Range* 3. *At all frequencies the results are derived from the simplified formulae* (*used in preceding paragraphs*) *only if the first decades of the Bridge are at zero* (*see* '*Discrimination and Accuracy*'—*next section*)

DISCRIMINATION AND ACCURACY

The Low Impedance Adaptor serves merely as an impedance invertor between the Bridge and the Unknown, measurement being made on the Bridge itself. This has two decade switches and a continuously-variable control on each component and is consequently capable of a discrimination of 0·02 per cent. Such a high discrimination extended to the Adaptor ranges can be useful for comparison of component values or the detection of small changes such as in temperature coefficient measurements. This order of discrimination is obtained only when the first decades are in use and, in these circumstances, the approximations referred to under 'Principle of Operation' are not justified. Absolute values cannot be derived with an accuracy of 1 per cent, therefore, using the simplified formulae, but can be calculated from the balance equation:

$$R^2/Z_u + 2R = 1/(G_b + j\omega C_b)$$

where R^2 is 10^4, 10^5, 10^6 and 10^7 ohms for Q221A Ranges 1-4, respectively.

It may be found more convenient to make an absolute measurement without using the first decades and employ this as a reference for subsequent comparisons where increased discrimination is important.

From the foregoing considerations it follows that Range 4 of the Adaptor should be used for measurements up to only 10 ohms if maximum absolute accuracy is required. Above this value the Bridge should be used. Where high discrimination between values of resistance is essential and the first decades are used, the true value can be found either from the balance equation or by reference to the correction curves given on page 36.

If an unstable value of impedance, such as the resistance of a defective switch contact, is to be measured, only an approximate reading is necessary. The continuously-variable controls alone should be used to obtain balance. If the switched decades are operated, the discrimination may be such that the random drift in resistance occurs more rapidly than the controls can be adjusted, making balance impossible.

CURRENT THROUGH THE UNKNOWN

The absolute value of alternating current passing through the Unknown cannot be specified accurately: the Bridge measurements are obtained by comparison. The values given are, therefore, approximate.

Adaptor Range	R.M.S. Current through Unknown
1	8mA
2	5mA
3	2mA
4	0·7mA

PRINCIPLE OF OPERATION

Refer to Fig. 25. When the Bridge is balanced, the voltage across the current transformer winding is, for all practical purposes, zero. Resistor R_2 is therefore effectively in parallel with the Unknown, Z_u. If the value of R_2 is very high compared with Z_u, the shunting effect is negligible and can be ignored. The current through the Unknown is then equal to $E/(R_1+Z_u)$. If R_1 also is made very high compared with Z_u, the effect of Z_u on the current is negligible and can be ignored. With these two approximations:

current through the Unknown,

$$i=E/R_1$$

voltage across the Unknown,

$$e=Z_u \cdot E/R_1$$

current fed back to Bridge,

$$I=e/R_2=Z_u \cdot E/R_1R_2.$$



Fig. 25. Principle of Operation

For convenience, R_1 and R_2 are made equal, so that:

$$R_1=R_2=R$$

and $$I=Z_u \cdot E/R^2$$

or $$Z_u=R^2 \cdot I/E \quad . \quad . \quad . \quad . \quad .(1)$$

The current I flows in the Bridge current transformer and E is the potential provided by the voltage transformer. The ratio I/E is therefore the admittance represented by the Bridge dial readings at balance. The value of the Unknown impedance is obtained (using equation 1) by multiplying the admittance read off the Bridge by R^2 and translating mhos as ohms. By making R^2 a suitable multiple of 10, the system becomes direct-reading.

It was assumed above that the value of the series resistors of the T-network was so high compared with the Unknown impedance that their shunting effect could be ignored. The balance equation is:

$$R^2/Z_u+2R=1/(G_b+j\omega C_b)$$

where G_b and C_b are the conductance and the capacitance values read from the Bridge dials. The approximations made in the derivation of equation (1) are based on the assumption that $|R^2/Z_u|$ is much greater than $2R$. This approximation is valid *provided that the first decades are not used.*

CIRCUIT DESCRIPTION

A circuit diagram is provided at the end of this Handbook. Four pairs of equal-value resistors form the switched series elements of a T-network. For most measurements the Blue and White plugs are effectively connected in parallel. The shunt arm of the network is then completed when the Unknown is connected between their inners and Neutral.

The series resistance of each arm is, for the four ranges, $\sqrt{10^4}$, $\sqrt{10^5}$, $\sqrt{10^6}$ and $\sqrt{10^7}$ ohms. The resistors in the Adaptor have slightly lower values to allow for the resistance of connecting leads and the Bridge transformer windings. The capacitors correct the small phase-shaft introduced by the leakage inductance of these transformers.

As stated in the Introduction, the function of the Adaptor is to reverse the method of measurement as compared with the Bridge. Thus the Blue plug of the Adaptor, although associated with the E plug of the Bridge, serves to feed a constant *current* to the Unknown. Similarly the White plug of the Adaptor feeds back a voltage o the Bridge.

MAINTENANCE AND SERVICE

A circuit diagram of the Adaptor is provided at the back of this Handbook. The Bridge itself may be used to check the values of the Adaptor components. The switch and all plugs and sockets must be maintained in good order.



CAPACITANCE CONVERSION CHART


Correction curves for Low Impedance Adaptor Q 221
Measurement of Resistance only
% Correction to be added
%
Range 2
Range 3
Range 4
7
6
5
4
3
2
1
0
7%
6
5
4
3
2
1
0
0·1
1
10
100
OHMS
RESISTANCE VALUE MEASURED
CORRECTION CURVES (RESISTANCE)

PARALLEL TO SERIES CONVERSIONS

Any impedance (or admittance) can be expressed in two ways:

(a) parallel components of the in-phase and quadrature terms

(b) series components of the in-phase and quadrature terms

Given values for the components in one form — (a) or (b) — values for the alternative form can be calculated to give an exact equivalent circuit at any one specified frequency of operation.

Because the Standards of the B221A Bridge are in parallel, the Unknown is measured in form (a). Expressions for obtaining the equivalent series components are given below. In the case of Low Impedance measurements, the bridge readings are in form (b).

$$R_s = 1/[G(1 + Q^2)]$$

$$C_s = C[1 + (1/Q^2)]$$

$$L_s = -1/[\omega^2 C\,(1 + 1/Q^2)]$$

$$X_s = X/[1 + (1/Q^2)]$$

$[Q = 2\pi fC/G]$

At 1592 Hz, $2\pi f = 10^4$.

where G and C are the values of the parallel components obtained from the bridge.

If $Q \gg 1$:

$R_s \simeq 1/Q^2G \quad C_s \simeq C \quad L_s \simeq 1/(\omega^2 C) \quad X_s \simeq X$

If $Q \ll 1$:

$R_s \simeq 1/G \quad C_s \simeq C/Q^2 \quad L_s \simeq Q^2/(\omega^2 C) \quad X_s \simeq Q^2X$

Note that Q can also be derived directly from the Bridge display (see 'Interpretation of Results' p. 11).

DECADE EQUIVALENTS

G	·001μmho	·01μmho	0·1μmho	1μmho	10μmho	100μmho	1m.mho	10m.mho	100m.mho
R	1000MΩ	100MΩ	10MΩ	1MΩ	100kΩ	10kΩ	1kΩ	100Ω	10Ω

C	0·1μμF	1μμF	10μμF	100μμF	·001μF	·01μF	0·1μF	1μF	10μF
L●	100kH	10kH	1000H	100H	10H	1H	100mH	10mH	1mH

● at 1592 Hz

WAYNE KERR
G
+ –
Cal
Zero
Off Volts On
+ –

Part 3 - Autobalance Adaptor AA221

INTRODUCTION

Autobalance Adaptor AA221 enables the Universal Bridge B221A, and Low Impedance Adaptor Q221A, to be operated without any manual balancing procedures. By suitable choice of Bridge Range, the full value of the Unknown can be displayed on the two meters or—with the Bridge decades set-up for the first two digits of a reading—the meters complete a four-figure read-out. In either case the B221A vernier controls are set to 0 and a continuous balance is maintained electronically.

Outputs are available, at a rear sub-panel, for operating recorders, pass/reject mechanisms, alarm systems, remote meters, control circuits or digital voltmeters with their associated print-out facilities. Thus the B221A/AA221 combination is ideal for continuous measurements of changing impedance/admittance values, such as the determination of temperature coefficients, and the rapid batching of components on production lines or in Goods Inwards departments.

The Adaptor includes a source and detector operating at 1592 Hz and the transistor circuits used are powered from an internal battery. Thus the Bridge and Adaptor can be used even when a.c. supplies are not available. When operation from 110 or 200–250V a.c. is desired, Power Supply Unit PS109, which is available as an optional extra, can be substituted in place of the battery.

All the advantages associated with the Bridge are retained. For instance, long measurement leads can be used and components can be measured whilst still connected in a circuit. The ancillaries available for the Bridge—Low Impedance Adaptor Q221A and the various conductivity and permittivity cells—can be employed whilst the AA221 is in use.

The accuracy and discrimination available from B221A and Q221A are fully maintained in all respects when the Autobalance Adaptor is employed. The basic (x1) measurement ranges are unaffected but the extension provided by the 'C 0.1' switch is not available while the AA221 is in use. The Adaptor is housed in a case of the same size and style as that of the Bridge itself.

SPECIFICATION

Measurement Ranges Accuracy, Discrimination	See Bridge Specification (Page 9) and Low Impedance Adaptor Specification (Page 28).
Source Frequency	1592 Hz ($\omega = 10^4$) $\pm 1\%$.
Meter Display	Simultaneous readings of conductance/resistance and capacitance/inductance. Linear mirrored scales, $4\frac{1}{4}$ in.
G and C Outputs	Voltmeter: 0 – 100mV (output impedance 200kΩ, balanced). Record: 100μA ($>$ 10Ω affects meters by $< 1\%$).
Power Requirement	Internal 9-volt battery type PP9 or equivalent. Power Supply Unit PS109 (optional extra) can be substituted in place of the battery. Operates from 110 or 200 – 250V, 50 – 60 Hz.
Dimensions	Width: 17 in. (43 cm.). Height: $11\frac{1}{2}$ in. (29 cm.). Depth: $7\frac{1}{2}$ in. (19 cm.).
Weight (including battery)	$19\frac{1}{4}$ lb. (8.7 kg.).

OPERATING INSTRUCTIONS - AA221 WITH B221A ONLY

Note: Certain Bridges must be modified slightly before they are used with the Autobalance Adaptor. The instruments affected, and details of the modification, are shown in smaller type on page 12.

SETTING UP

1 Disconnect the Universal Bridge B221A from the a.c. power supply.

2 Use the jack leads supplied to connect the Adaptor AA221 to the source and detector sockets on the back of the B221A Bridge. Turn the Range switch of the Bridge to position 7 and set all switched decades and both vernier controls to zero.

3 Switch the Adaptor to Volts. If the meter deflection lies on the coloured sector of the scale, the battery condition is satisfactory.

4 Set the +/— switches for G and C on the Bridge to the position required for measurement. Set the G switch on the Adaptor to the opposite sign to that of the Bridge and the C switch on the Adaptor to the same sign as that of the Bridge. Switch the Adaptor to "On".

5 Link together the two neutrals of the Bridge measurement leads but leave the "inners" open circuit. Adjust the Trim G and Trim C Bridge controls for zero reading on the corresponding Adaptor meters.

6 Switch in one minor decade step of G (on Range 7 this will represent a reading of 1 mmho). Adjust 'Cal F.s.d.' to give full-scale deflection on the G meter and adjust 'Cal Zero' to give zero deflection on the C meter.

7 Switch out the minor decade step, and return the G +/— switch on the Adaptor to the setting corresponding with that of the Bridge.

Since the adjustments described in step 6 are not completely free from interaction, some repetition may be necessary if a large alteration is made to either control.

MEASUREMENT PROCEDURE

1 When the trimming and calibration checks (above) are completed, connect the Bridge measurement leads to the Unknown.

2 If, on range 7, neither meter reads more than 10% of full-scale deflection, progressively reduce the Range switch setting until such deflection is obtained. The meter reading(s) will give a good indication of the value of the Unknown but greater reading accuracy can be obtained by transferring the first digit of the reading to the second switched decade and reducing the Range switch setting by 1. Even greater accuracy is obtained by transferring the first and second digits of the original meter reading to the first and second switched decades, respectively, and turning the Range switch down by 2.

3 Should the meter showing the minor term of the Unknown read more than 10% when the Bridge range has been reduced by 1 from the original setting, the first digit of the minor reading must be transferred to the appropriate second switched decade before the Range switch setting is reduced further.

4 If the meter deflection exceeds full-scale on Range 7, the first and second decades must be adjusted as necessary, on this range, to bring the meter deflection(s) on scale.

5 The first two figures of the value of the Unknown are given by the switched Bridge decades and the third and fourth figures by the corresponding meter readings from the Adaptor. The Adaptor +/— (G and C) switches must be set to positions corresponding with those of the Bridge. If set to opposite positions the meter readings would have to be subtracted from (rather than added to) the decade settings.

6 If it is desired to measure variations in either sense about a particular value, the vernier control(s) can be used to off-set the meter reading(s) to any desired extent. The vernier controls must be returned to zero when absolute measurements are to be made.

IMPORTANT

Unless the Bridge controls are so adjusted that both meter readings lie on-scale, neither reading can be regarded as reliable.

MEASUREMENT CONDITIONS

Approximate values of the voltage appearing across the Unknown at balance, when the Bridge is operated with the Autobalance Adaptor, are as follows:

Range	Volts (r.m.s.)
1	25 (±20 per cent)
2 and 3	2·5 „
4 and 5	250mV „
6 and 7	25mV „

The following resistive loading produces less than 1% change in the reading of f.s.d.

Range 1	100kΩ	Between inner and braid of 'source' (E) lead.
„ 2, 3	1kΩ	
„ 4, 5	10Ω	
„ 6, 7	0·1Ω	

Range 1, 2	250kΩ	Between inner and braid of 'detector' (I) lead.
„ 3, 4	2·5kΩ	
„ 5, 6	25Ω	
„ 7	0·25Ω	

Capacitive loading between inner and braid of the E lead causes a frequency change in the oscillator. With either meter at full-scale deflection, the effect will cause the other meter to read not more than 1 per cent of f.s.d. for the following loadings:

Range 1	200pF
„ 2, 3	0·02μF
„ 4, 5	2μF
„ 6, 7	20μF

Capacitive loading between inner and braid of the *I* lead changes the effective input impedance of the detector, causing a phase shift in the feedback current. With either meter at full-scale deflection, the effect will cause the other meter to read not more than 1 per cent of f.s.d. for the following loadings:

Range 1, 2	500pF
„ 3, 4	0·05μF
„ 5, 6	5μF
„ 7	50μF

Note: The effect of shunt loading can be minimised by making the f.s.d. and zero adjustments with such loading applied. This may not be possible with 3-terminal in-situ measurements but, if necessary, such loading could be measured separately and simulated for the trimming operations.

MONITORING OUTPUTS

A three-position switch on the rear sub-panel can be set for the following conditions:

1 *NORMAL*

G and C jack sockets disconnected from internal circuits.

2 *VOLTMETER*

G and C jack sockets provide 0 – 100 mV. Output impedance 200kΩ, balanced. Neither lead of this output should be grounded. Normally the f.s.d. and zero adjustments will be made with the external voltmeter(s) connected. When agreement must be obtained between the internal and external meters, the pre-set controls (RV5, RV6) can be adjusted. The procedure is described on page 47.

3 *RECORD*

G and C sockets provide 100μA. Recorders of resistance 10 ohms or less affect meter reading by less than 1 per cent. As with 'Voltmeter', f.s.d. and zero adjustments will normally be made to suit the external recorder.

The G and C meters of the Adaptor remain operative on all three switch positions.

OPERATING INSTRUCTIONS - AA221 WITH B221A & Q221A

Setting-up, and the measurement procedure, are essentially a combination of the instructions given for the AA221 (with B221A) and those for the Q221A, with the proviso that the Low Impedance Adaptor should be on the 0 – 100Ω range (as shown on the Q221A range plate) when the AA221 is first switched on. This procedure ensures that the load imposed by Q221A is initially at a minimum (maximum resistance) and enables the oscillations of the AA221 source to build up and stabilise in the shortest time (approx. 30 seconds). For ease of reference, the appropriate sequence of setting-up operations is repeated here in full.

SETTING UP

1 Disconnect the Universal Bridge B221A from the a.c. power supply.

2 Use the jack leads supplied to connect the Adaptor AA221 to the source and detector sockets on the back of the B221A Bridge. Turn the Range switch of the Bridge to position 3 and set all switched decades and both vernier controls to zero.

3 Remove the measurement cables from the Bridge and connect them to the sockets provided on the Low Impedance Adaptor.

4 Connect the two cables from the Adaptor to the corresponding sockets on the Bridge.

5 Turn the Low Impedance Adaptor to the 0–100Ω range (as shown on the Q221A range plate) and connect the two leads from the 'blue' socket to either side of the impedance to be measured.

6 Connect the neutral (green) lead from the 'white' cable between the Unknown and the neutral clip from the 'blue' cable.

7 Clip the 'inner' (red) lead from the 'white' cable on to the neutral clip of this same cable. This provides the necessary condition of only a single-point contact between one side of the Unknown and the detector for the initial trimming adjustment.

8 Switch the AA221 to 'Volts'. If the meter deflection lies on the coloured sector of the scale, the battery condition is satisfactory.

9 Set the +/— switches on the Bridge and on the Adaptor as follows:
both at 1 for R and L,
G at 1 and C at —1 for R and C,
either or both (G and C) at –1 for mutual or transfer impedance.

Switch the AA221 Adaptor to 'On' and wait one minute for the oscillator to stabilise.

10 Turn the Low Impedance Adaptor to the desired range and adjust the Trim G and Trim C Bridge controls for zero reading on the corresponding Adaptor meters.

11 Turn the Adaptor G switch to the *opposite* polarity to that of the Bridge G switch.

12 Switch in one minor decade step of G (on Range 3 this will represent a Bridge dial reading of 0·1 μmho). Adjust 'Cal.F.s.d.' to give full-scale deflection on the G meter and adjust 'Cal. Zero' to give zero deflection on the C meter.

13 Switch out the minor decade step, and return the G +/— switch on the Adaptor to the setting corresponding with that of the Bridge.

Since the adjustments described in step 12 are not completely free from interaction, some repetition may be necessary if a large alteration is made to either control.

MEASUREMENT PROCEDURE

1 Check the trimming adjustment as described in the previous section ('Setting Up'). Transfer the red lead from the 'white' cable to the other side of the Unknown. [This operation is identical to that described in detail on page 29, step 8, and the correct measurement connections are shown in the right-hand illustration of Fig. 21].

2 Adjust the second switched decades of the Bridge to obtain on-scale deflections on both meters. The first switched decades should not be employed if the measured value is to be derived from the simplified expressions shown for each of the four ranges on the Low Impedance Adaptor. The Bridge must remain on Range 3.

3 The equivalent series components of the Unknown are derived as stated on page 30, except that the meter readings replace those of the vernier controls.

MEASUREMENT CONDITIONS

The absolute value of alternating current passing through the Unknown cannot be specified accurately: the Bridge measurements are obtained by comparison. The values given are, therefore, approximate.

Adaptor Range (as on Q221A plate)	R.M.S. Current through Unknown (B221A/Q221A with AA221)
0—·1Ω	25
0—1Ω	8
0—10Ω	2·5
0—100Ω	0·8

SPECIALISED MEASUREMENTS

The specialised measurement techniques described for the B221A and the Q221A apply when the Autobalance Adaptor is in use provided that the following points of difference are borne in mind.

1 The two vernier controls of B221A are replaced by the corresponding AA221 meters.

2 The source of AA221, set at 1592 Hz, operates at a slightly lower level than that of B221A.

3 Provided only that the appropriate Range has been selected on B221A (and Q221A when in use), and the Bridge decades (when required) set to provide on-scale deflections on both meters, the Bridge balance is maintained electronically.

POWER SUPPLY UNIT PS109

Power Supply Unit PS109 can be fitted into the AA221 in place of the battery holder, providing for operation from a.c. supplies of 110V or 200-250V, 50-60 Hz. The Unit produces a stabilised d.c. output of 9 volts. Before connecting the PS109 to the a.c. supply, check that the primary taps on T1 are connected as follows:

230-volt supply
2-3 linked (no other connection)
1 live
4 neutral

110-volt supply
1-2 linked, to neutral
3-4 linked, to live

In addition to the two line connections, as above, the green wire of the power cable must be connected to a ground terminal. The fuse (FS1) should be 25mA for 110-volt operation or 10mA for 230-volt operation.

One lead to the primary is switched by the Off/Volts/On control. A bridge rectifier is energised from the secondary of T1 and a capacitance-input R-C smoothing circuit is employed (see Fig. 26). The emitter of VT4 is held at a reference potential by zener diode MR1. Any variation in the h.t. line potential causes a corresponding change in the potential of VT4 base (the mean operating point being established by adjustment of the preset control RV1). The change is amplified by VT4 and VT3 and applied, via the emitter follower VT2, to the base of the series regulator transistor VT1. The net effect of the circuit is to minimise any fluctuations in the h.t. potential.



Fig. 26. PS109 Circuit Diagram

PRINCIPLE OF OPERATION

A block diagram of the Autobalance Adaptor and its connections with the B221A Bridge is shown below. The Source oscillator circuit includes a thermistor bridge to provide an amplitude-stabilised signal at 1592 Hz ($\omega = 10^4$). The signal is applied to the primary of the B221A voltage transformer, and to the two phase-sensitive detectors (PSD's).

Any unbalance in the B221A Bridge circuit will result in a current flowing in the secondary of the current transformer. Negative feedback from the amplifier produces a current in opposition to this, so reducing the net core flux effectively to zero.

The voltage developed across the feed-back resistor (R_f) is applied to two phase-sensitive detectors, one of which is preceded by a 90° phase-shifting circuit. The detectors resolve the voltage into two components, proportional to the in-phase and quadrature terms, respectively, of the Unknown. The two values are indicated on the G and C meters.



Fig. 27. AA221 Block Diagram

CIRCUIT DESCRIPTION

A circuit diagram of the Autobalance Adaptor is provided at the back of this handbook.

SOURCE

Transistors VT1, VT2 and VT3 are employed in a positive feedback amplifier oscillating at 1592 Hz. The amplifying stage VT1 is coupled directly to the emitter-follower circuit of VT2, which provides the drive for the output stage (VT3). Transformer T1 and capacitor C5 are the components determining the frequency of oscillation. The amplitude of oscillation is stabilised by the bridge circuit consisting of the centre-tapped winding (6-7-8) on T1, thermistor TH1 and resistor R8, these components automatically controlling the amount of positive feedback to minimise any changes in the output level.

Neon N1, fed from a multi-turn winding on the tuned transformer, serves as the visual reminder when the instrument is switched 'On'. The source output is taken from a tapping on the transformer to the 'Voltage' jack socket. Reference signals for the two phase-sensitive detectors are taken from the balanced winding (6-7-8) on T1.

AMPLIFIER

Input to the high-gain tuned amplifying stage VT4 is derived from the secondary of the B221A current transformer. Components C8 and L3 form the tuned circuit: L3 is connected to provide a voltage step-up and phase reversal. To ensure that performance of the tuned circuit is not impaired by heavy loading, an emitter-follower stage (VT5) is introduced before the next amplifying stage, VT6. Diodes MR1 and MR2 limit the extent of any overloading, which might otherwise introduce phase-shifts giving false on-scale readings. To simplify the provision of direct coupling, the transistor employed for the amplifying stage VT6 is of the NPN type.

Detector amplifier VT7 serves a dual purpose. Feedback is produced at the emitter and is taken from this point (by C12) to the feedback network. But the emitter circuit resistance can be altered so that a second output, taken from VT7 collector, can be adjusted. This second output is applied to the emitter-follower VT8 whose function is to provide a low output impedance for

feeding the phase-sensitive detector (VT9, VT10) and the phase-shifting circuits associated with VT11.

FEEDBACK NETWORK

This part of the circuit, represented by R_f in the Block Diagram, actually comprises:

a 100:1 potential divider (R30 and R29);

a fixed phase shift introduced by C17;

a variable phase shift (RV3/R31 with C16);

compensation for ambient temperature variations (TH2/R28); and

the feedback resistor R27.

Feedback (from R27 and C16) is applied to the first stage (VT4) of the amplifier. Also, the flux in the core of the B221A current transformer is maintained virtually at zero by the opposing current fed back from the amplifier. Since the current transformer secondary is closely coupled to the primary, the effect is the same as that of providing a balancing current in a primary winding.

The purpose of the 100:1 potential divider is simply to increase the effective value of the feedback resistor (R27) from 39kΩ to 3·9MΩ without the need for using high-value resistors with their associated shunt capacity difficulties.

Phase shift is required for Set Zero purposes to offset the phase-shift occurring in the overall Bridge circuits (including the primary-to-secondary characteristics of the B221A voltage and current transformers). The set phase-shift of approximately 10° introduced on the main feedback line by C17 enables the control provided by RV3/C16 to become effective in either sense about zero.

Variations in the ambient temperature could (by affecting TH1) cause small changes in the amplitude of the source signal. The effect of these is nullified by thermistor TH2, which produces a compensating change in the amplifier sensitivity.

PHASE-SENSITIVE DETECTOR (G)

The in-phase component of the feedback, whose magnitude is determined by the value of the G term of the Unknown, is selected by the circuit associated with VT9 and VT10. Gating signals for these transistors are provided by winding 6-7-8 on T1 of the source circuit. The unidirectional output is applied, via sign-reversing and other switching arrangements, to the G meter.

90° CIRCUIT

An output from VT8 is applied to the network associated with VT11. Components R32 and C19, with VT11, form the essentials of an operational amplifier producing a phase shift approaching 90°. The small additional phase-shift is produced by components RV4 and C18 preceding the operational amplifier. Thus the output from VT11 is proportional to the feedback but is in quadrature with it. Emitter-follower VT12 provides a low-impedance output for driving the phase-sensitive detector VT13, VT14.

PHASE-SENSITIVE DETECTOR (C)

Operation of this detector is similar to that of the 'G' detector but, since its input has been shifted in phase by 90°, it responds to the quadrature component of the feedback, whose magnitude is determined by the value of the C/L term of the Unknown.

VOLTMETER/RECORD OUTPUTS

The unidirectional outputs from the two phase-sensitive detectors, G and C respectively, are applied to two identical metering circuits. Each meter is a 100μA movement and is shunted by the series combination of a fixed and pre-set resistor. When the rear-panel switch is set to NORMAL, the jacks are isolated from the Adaptor circuits.

On VOLTMETER the potential developed across the fixed resistor is available at the jack. The value of this, normally 100mV at full-scale deflection on the Adaptor meters, can be varied by adjustment of the pre-set controls.

On RECORD the internal meter circuit is interrupted by the jack. The external recorder should have a resistance of less than 10 ohms if the Adaptor meter readings are to be affected by less than 1% f.s.d.

MAINTENANCE AND SERVICE

SOURCE

The output can be checked at the voltage (E) jack socket JK1. The frequency should be 1592 Hz ±1%. If outside this tolerance, adjust the core of transformer T1. The amplitude should be 70V peak-to-peak ±20%. If low, the fault should be localised by checking voltages against the typical values provided on this page.

AMPLIFIER/DETECTOR

1 Disconnect the negative terminal of C12 from VT7.

2 Adjust an audio oscillator to provide an output of 1mV r.m.s. at 1592 Hz. Apply this, using a 3·9-MΩ resistor in series with the tip connection, to the Detector jack, JK2.

3 Connect an a.c. valve voltmeter between chassis and C14 —ve (VT8 emitter).

4 Set 'Cal.F.s.d.' fully clockwise and, using a trimming tool, adjust L3 for maximum reading on the voltmeter. This should be at least 35mV r.m.s. If less than this, the fault should be localised by checking voltages against the typical values provided on this page.

5 Replace the connection between C12 —ve and VT7

The following procedure is particularly relevant if difficulty has been experienced in making the calibration adjustments for 'F.s.d.' and 'Zero' (step 6 on page 41).

6 Make the appropriate 'Source' and 'Detector' connections between the AA221 and a B221A Bridge (disconnected from the a.c. supply). Set all the Bridge decade and vernier controls at zero, and turn the Range switch to position 7.

7 Set the +/— switches for G and C, on Bridge and Adaptor, to '1' (+).

8 Adjust Trim C and Trim G on the Bridge for zero readings on the corresponding meters of the Adaptor.

9 On the Adaptor only, set the C switch to —1. Switch in 1 minor decade of C on the Bridge.

10 Set 'Cal. Zero' fully clockwise and adjust 'Cal.F.s.d.' for full-scale deflection on the C meter.

11 Using a trimming tool, adjust the core of L3 to obtain a reading of 4 minor divisions on the G meter.

12 On the Adaptor only, set the G switch to '—' and adjust 'Cal. Zero' for zero reading on the G meter.

13 Re-set the minor decade C switch at zero and check that the Bridge remains balanced.

14 Switch in 1 minor decade of G on the Bridge.

15 Adjust the pre-set potentiometer RV2 (in AA221) for full-scale deflection on the G meter, and RV4 for zero reading on the C meter.

Note: It is important to check that the Bridge remains trimmed during the above adjustments.

VOLTMETER OUTPUT

Zero the Bridge in the normal manner and obtain full-scale deflection on the C meter by switching-in one minor C decade and setting the C switches on Bridge and Adaptor to opposite signs. Connect a d.c. Valve Voltmeter to the C output jack (JK3) and switch to 'Voltmeter'. Adjust RV5 for full-scale deflection (100mV) on the voltmeter.

A similar procedure should be carried out for setting-up RV6 on the G circuit, taking the output from JK4.

Typical Voltages (D.C.)

Figures quoted are a guide and variations of ±20% are normal. Avometer model 8 (+ve to chassis).

Transistor	Base	Emitter	Collector
VT1	1·8	1·7	3·5
VT2	3·5	3·5	8
VT3	0·79	0·66	8
VT4	2	1·3	8
VT5	8	6·7	8
VT6	6·7	6·6	2·2
VT7	2·2	2·3	5·5
VT8	5·5	5·5	8
VT9	0·35	-	-
VT10	0·35	-	-
VT11	1	1	4
VT12	4	4·1	8
VT13	0·35	-	-
VT14	0·35	-	-





Ref.	Value	Type
R1	98·6 Ω	± 0·5 per cent
R2	314·8 Ω	± 0·5 per cent
R3	998·5 Ω	± 0·5 per cent
R4	3·161k Ω	± 0·5 per cent
R5	98·6 Ω	± 0·5 per cent
R6	314·8 Ω	± 0·5 per cent
R7	998·5 Ω	± 0·5 per cent
R8	3·161k Ω	± 0·5 per cent
C1	6000μμF	± 5 per cent
C2	450μμF	± 5 per cent
C3	68μμF	± 5 per cent
C4	10μμF	± 1μμF

(Above) — B221A Circuit Diagram
(D3125, Issue AA)

Left — Q221A Components List and Circuit Diagram
(D9678, Issue F)

B221A COMPONENTS LIST

Ref.	Value	Tolerance (per cent)
R1–R4	Not used	
R5	5·6 kΩ	±10
R6	470 kΩ	±10
R7	Not used	
R8	22 kΩ	±10
R9	10 kΩ ⎫ Two in	±10
R9	10 kΩ ⎭ parallel	±10
R10	1·5 kΩ	±10
R11	100 kΩ	±10
R12	100 kΩ	±10
R13	100 kΩ	±10
R14	3·3 kΩ	±10
R15	100 kΩ	±0·05*
R16	10 kΩ	±0·05*
R17	390 kΩ	±1
R18	1 MΩ	±1
R19	1 kΩ	±10
R20	100 kΩ	±10
R21	1 kΩ	±10
R22	2·2 MΩ	±10
R23	22 kΩ	±10
R24–R31	Not used	
R32	2·2 MΩ	±20
R33	1 MΩ	±10
R34	1 MΩ	±10
R35	1 MΩ	±10
R36	1 MΩ	±10
R37	Not used	
R38	22 kΩ	±10
R39	A.I.C.	
R41	390 kΩ	±1
RV1	50 kΩ	Linear
RV2	10 kΩ	Lin. to 5 per cent
RV3	25 kΩ	Linear
RV4	1 MΩ	Log
RV5	50 kΩ	Linear
FS1	500 mA	

* Adjusted in circuit

Ref.	Value	Type
C1–C6	Not used	
C7	5000 μμF	±1 per cent
C8	0·01 μF	350 V
C9	8 μF	450 V
C10	0·01 μF	350 V
C11	440 μμF	Variable
C12	0·001 μF	+0, −1 per cent
C13	3—30 μμF	Trimmer
C14	0·01 μF	+0, −1 per cent
C15	22 μμF	±10 per cent
C16	3—30 μμF	Trimmer
C17	100 μμF	Variable
C18	33 μμF	±10 per cent
C19	16 μF	275 V
C20	5000 μμF	±1 per cent
C21	0·01 μF	350 V
C22	5000 μμF	±1 per cent
C23	0·001 μF	350 V
C24	8 μF	450 V
C25–C33	Not used	
C34	0·001 μF	
C35	0·001 μF	
C36	0·5 —5 μμF	Trimmer
C37	See page 25	
C38	0·5 —5 μμF	Trimmer
C39	1 μF	70 V
C40	10 μμF	±10 per cent
C41	8 μF	
T1	Voltage Transformer Type R.T. 3·8	
T2	Current Transformer Type R.T. 4·8	
T3	Power Transformer Pri: 0–100/125 and 200/250V Sec: 210–0–210 V, 40 mA 6·3 V, 1·75 A 6·3 V, 1·0 A	

§Measurement Cable (2 per inst.)	D10065
†Low Capacity Clip leads LCC3	D10642
Power Cable Assembly (CT530 only)	D10542
Lid Assembly (CT530 only)	D10544
Bridge Cased Assembly (CT530 only)	D10594

§Specify BNC (bayonet) or Miniature Pye (threaded) connectors.
†LCC3 is fitted with BNC connectors.
LCC2 is fitted with Min. Pye connectors.



SWITCH	POSITION	FUNCTION
1	1	C +
	2	C −
2	1	G +
	2	G −
3	1	OFF
	2	CHECK SUPPLY
	3	OPERATE
4	1	NORMAL
	2	VOLTMETER
	3	RECORD

AA221 Circuit Diagram (D7217, Issue V)

www.manualsplus.com
1003879

THE WAYNE KERR COMPANY LIMITED

WAYNE KERR NEW MALDEN SURREY, ENGLAND Cables: Waynkerr, New Malden

Telex: 262333 Telephone: 01-942 2202

WAYNE KERR CORPORATION 22 FRINK STREET, MONTCLAIR NEW JERSEY 07042 U.S.A.

Telephone: (201) 746-2438 Cables: Waynkerr, Montclair, N.J.

Printed in England TWP/7/67/500